

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : AR1268PR090320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 04 08:20:28 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047081.D 1600 =PR047083.D 800 =PR047082.D
 200 =PR047080.D 100 =PR047079.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.213	0.998	1.110	1.319	1.215	1.171	E6	10.39
2) SA	Decachlorobiphenyl	2.914	2.630	2.744	3.176	3.280	2.949	E6	9.38
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.367	2.239	2.259	2.588	2.609	2.412	E5	7.33
42) L9	AR-1268-2	2.209	2.084	2.114	2.406	2.399	2.242	E5	6.83
43) L9	AR-1268-3	1.869	1.757	1.790	2.042	2.064	1.904	E5	7.46
44) L9	AR-1268-4	8.913	7.987	8.336	9.595	9.469	8.860	E4	7.88
45) L9	AR-1268-5	6.437	6.235	6.291	6.794	6.891	6.530	E5	4.55

Signal #2 Calibration Files

400 =PR047081.D 1600 =PR047083.D 800 =PR047082.D
 200 =PR047080.D 100 =PR047079.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.042	3.721	3.822	4.428	4.322	4.067	E6	7.54
2) SA	Decachlorobiphenyl	2.291	2.244	2.249	2.383	2.364	2.306	E7	2.81
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : AR1268PR090320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 04 08:20:28 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047081.D 1600 =PR047083.D 800 =PR047082.D
 200 =PR047080.D 100 =PR047079.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.223	2.227	2.186	2.319	2.253	2.242 E6	2.19
42) L9	AR-1268-2	2.085	2.092	2.052	2.166	2.103	2.100 E6	1.98
43) L9	AR-1268-3	1.823	1.839	1.806	1.911	1.864	1.849 E6	2.22
44) L9	AR-1268-4	7.724	7.614	7.569	8.310	8.045	7.852 E5	4.03
45) L9	AR-1268-5	5.688	5.603	5.736	5.717	5.604	5.670 E6	1.10

(#) = Out of Range