

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090320\
 Data File : PR047064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2020 15:57
 Operator : DD\AJ
 Sample : MDL-AR1262-S-4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-S-4

Manual Integrations
 APPROVED

Ankita
 9/4/2020 2:40:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 07:31:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 14:22:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.767	3.895	20994622	72314778	19.076	18.317
2) SA Decachlor...	10.727	8.992	66012368	495.9E6	40.365	39.962
Target Compounds						
36) L8 AR-1262-1	8.175	7.000	3329603	8705607	31.770	30.317m
37) L8 AR-1262-2	8.758	7.543	5625570	38446261	28.666	23.624
38) L8 AR-1262-3	9.094	7.831	3883577	19595277	28.082	27.250
39) L8 AR-1262-4	9.187	7.895	2474563	31987513	23.785	26.050
40) L8 AR-1262-5	9.903	8.404	2113743	30043091	27.317	44.362 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090320\
Data File : PR047064.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2020 15:57
Operator : DD\AJ
Sample : MDL-AR1262-S-4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-S-4

Manual Integrations
APPROVED

Ankita
9/4/2020 2:40:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 07:31:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 14:22:08 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

