

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090320\
 Data File : PR047065.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2020 16:12
 Operator : DD\AJ
 Sample : MDL-AR1262-S-5
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-S-5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 07:32:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 14:22:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.768	3.896	20624422	72102501	18.740	18.263
2) SA Decachlor...	10.729	8.992	65484279	495.8E6	40.042	39.958
Target Compounds						
36) L8 AR-1262-1	8.174	7.002	3225239	8979392	30.774	31.270
37) L8 AR-1262-2	8.757	7.543	5459535	37584943	27.820	23.094
38) L8 AR-1262-3	9.093	7.831	3199054	16299831	23.133	22.667
39) L8 AR-1262-4	9.187	7.895	2276344	28756128	21.880	23.418
40) L8 AR-1262-5	9.903	8.404	2101406	16213282	27.158	23.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090320\
Data File : PR047065.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2020 16:12
Operator : DD\AJ
Sample : MDL-AR1262-S-5
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-S-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 07:32:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 14:22:08 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

