

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090420\
 Data File : PR047109.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2020 10:44
 Operator : DD\AJ
 Sample : MDL-AR1262-W-2
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-W-2

Manual Integrations
 APPROVED

Ankita
 9/4/2020 3:45:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 11:13:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 14:22:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.764	3.896	26317341	90928413	23.912	23.031
2) SA Decachlor...	10.722	8.992	73302773	563.4E6	44.823	45.405
Target Compounds						
36) L8 AR-1262-1	8.171	7.001	3438550	11765491	32.810	40.973
37) L8 AR-1262-2	8.753	7.544	5727897	44204235	29.187	27.162
38) L8 AR-1262-3	9.089	7.832	5037118	22471307	36.424m	31.249
39) L8 AR-1262-4	9.187	7.896	2897789	35061080	27.853	28.553
40) L8 AR-1262-5	9.901	8.405	2647055	18888754	34.209	27.892

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090420\
Data File : PR047109.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2020 10:44
Operator : DD\AJ
Sample : MDL-AR1262-W-2
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-W-2

Manual Integrations
APPROVED

Ankita
9/4/2020 3:45:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 11:13:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 14:22:08 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

