

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090420\
 Data File : PR047112.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2020 11:27
 Operator : DD\AJ
 Sample : MDL-AR1262-W-5
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1262-W-5

Manual Integrations
 APPROVED

Ankita
 9/4/2020 3:45:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 13:06:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 14:22:08 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.763	3.895	22014693	77437897	20.003	19.614
2) SA Decachlor...	10.721	8.991	66656809	517.5E6	40.759	41.702
Target Compounds						
36) L8 AR-1262-1	8.171	7.000	3014331	11639891	28.762	40.535 #
37) L8 AR-1262-2	8.752	7.543	5337317	39242721	27.197	24.113
38) L8 AR-1262-3	9.087	7.830	4558714	19119391	32.964m	26.588
39) L8 AR-1262-4	9.185	7.894	2681814	29431514	25.777	23.968
40) L8 AR-1262-5	9.900	8.403	1859300	17915890	24.029	26.455

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090420\
Data File : PR047112.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2020 11:27
Operator : DD\AJ
Sample : MDL-AR1262-W-5
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1262-W-5

Manual Integrations
APPROVED

Ankita
9/4/2020 3:45:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 13:06:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1262PR090320CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 14:22:08 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

