

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090518.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 06 03:46:35 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031774.D 1000 =PR031772.D 750 =PR031773.D
 250 =PR031775.D 50 =PR031776.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.398	1.944	2.292	2.308	2.123	2.213	E7	8.14
2) SA	Decachlorobiphenyl	2.581	2.727	2.459	2.555	2.682	2.601	E7	4.09
3) L1	AR-1016-1	5.936	4.763	5.545	5.923	5.535	5.540	E5	8.60
4) L1	AR-1016-2	5.835	4.740	5.486	5.792	5.223	5.415	E5	8.34
5) L1	AR-1016-3	8.956	7.299	8.356	8.913	9.420	8.589	E5	9.48
6) L1	AR-1016-4	7.025	5.812	6.612	7.078	7.186	6.742	E5	8.36
7) L1	AR-1016-5	6.451	5.274	6.017	6.446	6.553	6.148	E5	8.63
8) L2	AR-1221-1	2.685					2.685	E5	0.00
9) L2	AR-1221-2	2.061					2.061	E5	0.00
10) L2	AR-1221-3	6.276					6.276	E5	0.00
11) L3	AR-1232-1	4.898					4.898	E5	0.00
12) L3	AR-1232-2	2.248					2.248	E5	0.00
13) L3	AR-1232-3	7.758					7.758	E4	0.00
14) L3	AR-1232-4	7.529					7.529	E4	0.00
15) L3	AR-1232-5	4.279					4.279	E4	0.00
16) L4	AR-1242-1	2.529					2.529	E5	0.00
17) L4	AR-1242-2	4.787					4.787	E5	0.00
18) L4	AR-1242-3	2.226					2.226	E5	0.00
19) L4	AR-1242-4	1.724					1.724	E5	0.00
20) L4	AR-1242-5	5.993					5.993	E5	0.00
21) L5	AR-1248-1	7.881					7.881	E5	0.00
22) L5	AR-1248-2	8.611					8.611	E5	0.00
23) L5	AR-1248-3	1.118					1.118	E6	0.00
24) L5	AR-1248-4	1.043					1.043	E6	0.00
25) L5	AR-1248-5	6.495					6.495	E5	0.00
26) L6	AR-1254-1	1.894					1.894	E6	0.00
27) L6	AR-1254-2	1.151					1.151	E6	0.00
28) L6	AR-1254-3	2.080					2.080	E6	0.00
29) L6	AR-1254-4	1.657					1.657	E6	0.00
30) L6	AR-1254-5	1.198					1.198	E6	0.00
31) L7	AR-1260-1	1.317	1.165	1.251	1.277	1.296	1.261	E6	4.70
32) L7	AR-1260-2	1.700	1.562	1.625	1.636	1.685	1.642	E6	3.31
33) L7	AR-1260-3	1.960	1.848	1.890	1.870	1.918	1.897	E6	2.29
34) L7	AR-1260-4	1.373	1.332	1.323	1.327	1.338	1.338	E6	1.49
35) L7	AR-1260-5	3.051	3.120	2.989	2.881	2.931	2.995	E6	3.16
36) L8	AR-1262-1	8.435					8.435	E5	0.00
37) L8	AR-1262-2	8.108					8.108	E5	0.00
38) L8	AR-1262-3	6.947					6.947	E5	0.00
39) L8	AR-1262-4	1.854					1.854	E6	0.00
40) L8	AR-1262-5	1.315					1.315	E6	0.00
41) L9	AR-1268-1	7.220					7.220	E5	0.00
42) L9	AR-1268-2	9.749					9.749	E5	0.00
43) L9	AR-1268-3	4.137					4.137	E6	0.00
44) L9	AR-1268-4	3.279					3.279	E6	0.00
45) L9	AR-1268-5	9.491					9.491	E6	0.00

Signal #2 Calibration Files

500 =PR031774.D 1000 =PR031772.D 750 =PR031773.D
 250 =PR031775.D 50 =PR031776.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.586	3.716	4.360	4.433	3.852	4.189	E7	9.12
2) SA	Decachlorobiphenyl	1.906	1.901	1.785	1.945	2.324	1.972	E7	10.43
3) L1	AR-1016-1	1.151	0.893	1.073	1.155	1.160	1.086	E6	10.49
4) L1	AR-1016-2	1.744	1.423	1.643	1.712	1.815	1.667	E6	8.99
5) L1	AR-1016-3	3.015	2.375	2.833	2.959	2.968	2.830	E6	9.30

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090518.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 06 03:46:35 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031774.D 1000 =PR031772.D 750 =PR031773.D
 250 =PR031775.D 50 =PR031776.D

Compound			500	1000	750	250	50	Avg	%RSD

6)	L1	AR-1016-4	1.817	1.467	1.705	1.807	1.816	1.722 E6	8.72
7)	L1	AR-1016-5	1.644	1.342	1.527	1.644	1.771	1.586 E6	10.15
8)	L2	AR-1221-1	4.717					4.717 E5	0.00
9)	L2	AR-1221-2	3.541					3.541 E5	0.00
10)	L2	AR-1221-3	1.158					1.158 E6	0.00
11)	L3	AR-1232-1	4.991					4.991 E5	0.00
12)	L3	AR-1232-2	1.851					1.851 E5	0.00
13)	L3	AR-1232-3	5.425					5.425 E5	0.00
14)	L3	AR-1232-4	6.575					6.575 E5	0.00
15)	L3	AR-1232-5	1.016					1.016 E5	0.00
16)	L4	AR-1242-1	5.539					5.539 E5	0.00
17)	L4	AR-1242-2	1.200					1.200 E6	0.00
18)	L4	AR-1242-3	4.274					4.274 E5	0.00
19)	L4	AR-1242-4	3.746					3.746 E5	0.00
20)	L4	AR-1242-5	8.637					8.637 E5	0.00
21)	L5	AR-1248-1	1.764					1.764 E6	0.00
22)	L5	AR-1248-2	3.463					3.463 E6	0.00
23)	L5	AR-1248-3	2.462					2.462 E6	0.00
24)	L5	AR-1248-4	2.080					2.080 E6	0.00
25)	L5	AR-1248-5	1.513					1.513 E6	0.00
26)	L6	AR-1254-1	3.731					3.731 E6	0.00
27)	L6	AR-1254-2	2.833					2.833 E6	0.00
28)	L6	AR-1254-3	4.177					4.177 E6	0.00
29)	L6	AR-1254-4	1.951					1.951 E6	0.00
30)	L6	AR-1254-5	2.912					2.912 E6	0.00
31)	L7	AR-1260-1	3.060	2.648	2.835	3.052	3.338	2.987 E6	8.71
32)	L7	AR-1260-2	3.643	3.225	3.395	3.622	4.093	3.596 E6	9.10
33)	L7	AR-1260-3	2.891	2.553	2.695	2.903	3.200	2.849 E6	8.59
34)	L7	AR-1260-4	1.663	1.487	1.524	1.727	2.098	1.700 E6	14.31
35)	L7	AR-1260-5	3.393	3.150	3.147	3.540	4.308	3.508 E6	13.62
36)	L8	AR-1262-1	2.050					2.050 E6	0.00
37)	L8	AR-1262-2	1.442					1.442 E6	0.00
38)	L8	AR-1262-3	1.131					1.131 E6	0.00
39)	L8	AR-1262-4	1.526					1.526 E6	0.00
40)	L8	AR-1262-5	1.136					1.136 E6	0.00
41)	L9	AR-1268-1	1.228					1.228 E6	0.00
42)	L9	AR-1268-2	3.113					3.113 E6	0.00
43)	L9	AR-1268-3	2.801					2.801 E6	0.00
44)	L9	AR-1268-4	7.649					7.649 E5	0.00
45)	L9	AR-1268-5	7.476					7.476 E6	0.00

 (#) = Out of Range