

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 01:47:45 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047121.D 1600 =PR047123.D 800 =PR047122.D
 200 =PR047120.D 100 =PR047119.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.274	1.185	1.338	1.456	1.333	1.317	E6	7.52
2) SA	Decachlorobiphenyl	1.629	1.435	1.599	1.840	1.879	1.676	E6	10.95
3) L1	AR-1016-1	5.444	4.753	5.349	6.285	5.821	5.530	E4	10.31
4) L1	AR-1016-2	7.599	6.804	7.672	8.882	8.221	7.836	E4	9.87
5) L1	AR-1016-3	4.733	4.074	4.671	5.557	5.021	4.811	E4	11.24
6) L1	AR-1016-4	3.949	3.454	3.917	4.551	3.954	3.965	E4	9.83
7) L1	AR-1016-5	4.057	3.572	3.940	4.497	4.330	4.079	E4	8.80
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	7.927	6.912	7.607	9.079	9.218	8.149	E4	12.09
32) L7	AR-1260-2	0.957	0.854	0.934	1.117	1.146	1.001	E5	12.51
33) L7	AR-1260-3	7.737	6.782	7.476	8.616	8.760	7.874	E4	10.44
34) L7	AR-1260-4	0.889	0.808	0.885	1.012	1.022	0.923	E5	9.90
35) L7	AR-1260-5	1.769	1.703	1.813	1.990	1.970	1.849	E5	6.81
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR047121.D 1600 =PR047123.D 800 =PR047122.D
 200 =PR047120.D 100 =PR047119.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.292	3.980	4.470	4.982	4.704	4.486	E6	8.54
2) SA	Decachlorobiphenyl	1.247	1.205	1.288	1.382	1.347	1.294	E7	5.57
3) L1	AR-1016-1	1.548	1.385	1.564	1.797	1.659	1.591	E5	9.56
4) L1	AR-1016-2	2.624	2.393	2.697	3.017	2.811	2.708	E5	8.50
5) L1	AR-1016-3	1.564	1.417	1.588	1.806	1.657	1.606	E5	8.83

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 01:47:45 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047121.D 1600 =PR047123.D 800 =PR047122.D
 200 =PR047120.D 100 =PR047119.D

Compound			400	1600	800	200	100	Avg	%RSD	
6)	L1	AR-1016-4	8.598	7.571	8.532	9.748	8.765	8.643	E4	8.96
7)	L1	AR-1016-5	1.413	1.225	1.407	1.640	1.519	1.441	E5	10.64
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	2.874	2.547	2.799	3.283	3.512	3.003	E5	12.94
32)	L7	AR-1260-2	4.177	4.024	4.463	4.927	4.813	4.481	E5	8.72
33)	L7	AR-1260-3	3.886	3.668	3.970	4.380	4.129	4.007	E5	6.66
34)	L7	AR-1260-4	3.827	3.632	3.867	4.270	4.441	4.008	E5	8.37
35)	L7	AR-1260-5	1.558	1.562	1.631	1.693	1.678	1.624	E6	3.87
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

(#) = Out of Range