

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 06:16:11 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047151.D 1600 =PR047153.D 800 =PR047152.D
 200 =PR047150.D 100 =PR047149.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.267	1.216	1.196	1.468	1.279	1.285	E6	8.40
2) SA	Decachlorobiphenyl	1.711	1.428	1.541	1.884	1.936	1.700	E6	12.77
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	1.139	0.950	1.021	1.247	1.264	1.124	E5	12.24
37) L8	AR-1262-2	2.079	1.865	1.950	2.229	2.214	2.067	E5	7.74
38) L8	AR-1262-3	1.468	1.252	1.352	1.606	1.610	1.457	E5	10.79
39) L8	AR-1262-4	1.122	0.945	1.034	1.212	1.209	1.104	E5	10.44
40) L8	AR-1262-5	8.167	6.980	7.472	9.011	8.836	8.093	E4	10.74
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR047151.D 1600 =PR047153.D 800 =PR047152.D
 200 =PR047150.D 100 =PR047149.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.491	3.969	4.220	4.977	4.635	4.458	E6	8.67
2) SA	Decachlorobiphenyl	1.306	1.190	1.230	1.364	1.348	1.288	E7	5.83
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 06:16:11 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047151.D 1600 =PR047153.D 800 =PR047152.D
 200 =PR047150.D 100 =PR047149.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	3.070	2.681	2.843	3.338	3.301	3.047 E5	9.36
37) L8	AR-1262-2	1.852	1.659	1.770	1.926	1.907	1.823 E6	6.02
38) L8	AR-1262-3	7.658	6.857	7.207	8.002	7.782	7.501 E5	6.17
39) L8	AR-1262-4	1.318	1.210	1.255	1.356	1.319	1.292 E6	4.53
40) L8	AR-1262-5	7.327	6.588	6.801	7.604	7.515	7.167 E5	6.27
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range