

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 06:37:25 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047156.D 1600 =PR047158.D 800 =PR047157.D
 200 =PR047155.D 100 =PR047154.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.208	1.166	1.190	1.458	1.560	1.316	E6	13.68
2) SA	Decachlorobiphenyl	2.974	2.574	2.777	3.201	3.426	2.990	E6	11.25
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	2.472	2.193	2.347	2.656	2.882	2.510	E5	10.70
42) L9	AR-1268-2	2.303	2.031	2.173	2.449	2.632	2.318	E5	10.11
43) L9	AR-1268-3	1.930	1.701	1.827	2.072	2.285	1.963	E5	11.51
44) L9	AR-1268-4	0.911	0.778	0.850	0.985	1.044	0.913	E5	11.57
45) L9	AR-1268-5	6.577	6.049	6.377	6.824	7.144	6.594	E5	6.34

Signal #2 Calibration Files

400 =PR047156.D 1600 =PR047158.D 800 =PR047157.D
 200 =PR047155.D 100 =PR047154.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.475	3.884	4.141	5.124	5.512	4.627	E6	14.66
2) SA	Decachlorobiphenyl	2.363	2.214	2.307	2.437	2.498	2.364	E7	4.69
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090520CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Sep 05 06:37:25 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR047156.D 1600 =PR047158.D 800 =PR047157.D
 200 =PR047155.D 100 =PR047154.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.314	2.137	2.253	2.459	2.470	2.327 E6	6.06
42) L9	AR-1268-2	2.192	2.053	2.142	2.280	2.313	2.196 E6	4.79
43) L9	AR-1268-3	1.929	1.785	1.863	2.002	2.055	1.927 E6	5.60
44) L9	AR-1268-4	8.302	7.425	7.838	8.567	9.483	8.323 E5	9.40
45) L9	AR-1268-5	5.761	5.646	5.768	5.945	5.796	5.783 E6	1.85

(#) = Out of Range