

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090619CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 06 10:15:43 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040707.D 1600 =PR040709.D 800 =PR040708.D
 200 =PR040706.D 100 =PR040705.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.053	3.105	3.207	3.284	3.256	3.181	E6	3.12
2) SA	Decachlorobiphenyl	2.794	2.599	3.053	3.195	3.222	2.973	E6	9.04
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1	3.081	2.859	3.099	3.362	3.566	3.194	E4	8.59
9) L2	AR-1221-2	2.199	2.025	2.236	2.509	2.792	2.352	E4	12.79
10) L2	AR-1221-3	7.334	6.780	7.456	8.129	8.108	7.561	E4	7.52
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR040707.D 1600 =PR040709.D 800 =PR040708.D
 200 =PR040706.D 100 =PR040705.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.648	1.560	1.670	1.848	1.846	1.714	E7	7.45
2) SA	Decachlorobiphenyl	0.922	0.878	1.017	1.024	1.028	0.974	E7	7.13
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090619CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 06 10:15:43 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040707.D 1600 =PR040709.D 800 =PR040708.D
 200 =PR040706.D 100 =PR040705.D

	Compound	400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1	1.399	1.235	1.382	1.565	1.656	1.448 E5	11.41
9) L2	AR-1221-2	0.854	0.744	0.852	0.952	1.009	0.882 E5	11.58
10) L2	AR-1221-3	2.420	2.120	2.410	2.710	2.799	2.492 E5	10.84
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range