

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090619CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 06 09:58:11 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040712.D 1600 =PR040714.D 800 =PR040713.D
 200 =PR040711.D 100 =PR040710.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	3.124	2.948	3.065	3.286	3.089	3.103	E6	3.94
2)	SA	Decachlorobiphenyl	2.782	2.687	2.593	3.034	3.185	2.856	E6	8.63
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1	6.066	5.185	5.696	6.526	6.393	5.973	E4	9.13
12)	L3	AR-1232-2	3.056	2.752	2.891	3.342	2.867	2.982	E4	7.68
13)	L3	AR-1232-3	6.212	5.741	6.012	6.867	6.424	6.251	E4	6.83
14)	L3	AR-1232-4	3.036	2.798	2.931	3.429	2.981	3.035	E4	7.82
15)	L3	AR-1232-5	2.522	2.132	2.298	2.637	2.367	2.391	E4	8.20
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR040712.D 1600 =PR040714.D 800 =PR040713.D
 200 =PR040711.D 100 =PR040710.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.714	1.499	1.618	1.855	1.786	1.694	E7	8.28
2) SA	Decachlorobiphenyl	0.943	0.900	0.867	0.984	1.016	0.942	E7	6.42
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR090619CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Sep 06 09:58:11 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040712.D 1600 =PR040714.D 800 =PR040713.D
 200 =PR040711.D 100 =PR040710.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1	2.023	1.655	1.852	2.250	2.104	1.977 E5	11.65
12) L3	AR-1232-2	1.682	1.540	1.624	1.865	1.679	1.678 E5	7.10
13) L3	AR-1232-3	8.789	8.290	8.772	9.489	8.612	8.791 E4	4.99
14) L3	AR-1232-4	8.036	7.261	7.711	8.889	9.172	8.214 E4	9.75
15) L3	AR-1232-5	0.954	0.866	0.916	1.058	1.025	0.964 E5	8.14
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range