

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090619\
 Data File : PR040724.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2019 18:27
 Operator : SM\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 09:18:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 06 09:14:29 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.756	3.997	239.7E6	1213.7E6	78.169	74.158
2) SA Decachlor...	10.697	9.142	419.0E6	1420.5E6	150.737	152.256
Target Compounds						
21) L5 AR-1248-1	5.933	5.101	91742145	247.2E6	1486.290	1502.625
22) L5 AR-1248-2	6.206	5.340	125.4E6	334.5E6	1480.698	1495.387
23) L5 AR-1248-3	6.414	5.384	142.4E6	348.1E6	1493.355	1490.645
24) L5 AR-1248-4	6.818	5.559	165.1E6	455.0E6	1513.900	1490.554
25) L5 AR-1248-5	6.857	5.956	155.3E6	491.4E6	1514.686	1488.133

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090619\
Data File : PR040724.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2019 18:27
Operator : SM\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 09:18:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 06 09:14:29 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

