

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090619\
Data File : PR040710.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2019 15:04
Operator : SM\AJ
Sample : AR1232ICC100
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_R
LabSampleId :
AR1232101

Manual Integrations
APPROVED

Ankita
9/6/2019 4:26:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 10:11:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 06 09:58:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) AR-1232-1 (L3)
R.T. Response Conc
5.13 6392842 107.02
5.67 2867029 96.15
5.96 6423878 102.76
6.12 2981142 98.22
6.20 0 0

(11) AR-1232-1 #2 (L3)
R.T. Response Conc
4.38 21037296 106.43
5.12 16792556 100.06
5.30 8612037 97.97
5.38 9172229 111.67
5.56 10253831 106.39

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090619\
Data File : PR040710.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2019 15:04
Operator : SM\AJ
Sample : AR1232ICC100
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_R
LabSampleId :
AR1232101

Manual Integrations
APPROVED

Ankita
9/6/2019 4:26:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 10:11:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 06 09:58:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) AR-1232-1 (L3)
R.T. Response Conc
5.13 6392842 107.02
5.67 2867029 96.15
5.96 6423878 102.76
6.12 2981142 98.22
6.20 2367311 99.12

(11) AR-1232-1 #2 (L3)
R.T. Response Conc
4.38 21037296 106.43
5.12 16792556 100.06
5.30 8612037 97.97
5.38 9172229 111.67
5.56 10253831 106.39

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090619\
 Data File : PR040710.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2019 15:04
 Operator : SM\AJ
 Sample : AR1232IC100
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1232101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 10:11:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 06 09:58:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 9/6/2019 4:26:30 PM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	4.756	3.996	15447077	89278042	4.979	5.270
2) SA Decachlor...	10.699	9.142	31854030	101.6E6	11.152	10.788
Target Compounds						
11) L3 AR-1232-1	5.130	4.377	6392842	21037296	107.022	106.429
12) L3 AR-1232-2	5.666	5.122	2867029	16792556	96.152	100.064
13) L3 AR-1232-3	5.956	5.301	6423878	8612037	102.764	97.969
14) L3 AR-1232-4	6.119	5.385	2981142	9172229	98.223	111.667
15) L3 AR-1232-5	6.204	5.559	2367311	10253831	99.122m	106.391

A.J
 09/10/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.