

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090624\
 Data File : PR068489.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2024 20:19
 Operator : AJ\MA
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 02:42:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 07 02:32:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.648	2.982	28691227	136.7E6	25.566	26.433
2) SA Decachlor...	9.538	8.275	19885954	100.0E6	26.199	25.982
Target Compounds						
26) L6 AR-1254-1	5.776	4.950	11095354	70006781	268.115	267.347
27) L6 AR-1254-2	6.013	5.110	16976041	63196196	267.906	271.479
28) L6 AR-1254-3	6.405	5.535	15972026	91932903	264.035	266.513
29) L6 AR-1254-4	6.716	5.783	11827233	57079660	266.188	264.940
30) L6 AR-1254-5	7.170	6.231	13869209	83258996	265.147	263.210



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090624\
Data File : PR068489.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2024 20:19
Operator : AJ\MA
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 02:42:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 07 02:32:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 28691227
Conc: 25.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 2.982 min
Delta R.T.: 0.000 min
Response: 136670343
Conc: 26.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.538 min
Delta R.T.: 0.000 min
Response: 19885954
Conc: 26.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.275 min
Delta R.T.: -0.001 min
Response: 100028283
Conc: 25.98 ng/ml



#26 AR-1254-1


R.T.: 5.776 min
Delta R.T.: 0.000 min
Response: 11095354
Conc: 268.12 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 4.950 min
Delta R.T.: 0.000 min
Response: 70006781
Conc: 267.35 ng/ml




#27 AR-1254-2

R.T.: 6.013 min
Delta R.T.: 0.000 min
Response: 16976041
Conc: 267.91 ng/ml



#27 AR-1254-2


R.T.: 5.110 min
Delta R.T.: 0.000 min
Response: 63196196
Conc: 271.48 ng/ml



#28 AR-1254-3


R.T.: 6.405 min
Delta R.T.: 0.000 min
Response: 15972026
Conc: 264.03 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 5.535 min
Delta R.T.: 0.000 min
Response: 91932903
Conc: 266.51 ng/ml




#29 AR-1254-4

R.T.: 6.716 min
Delta R.T.: 0.000 min
Response: 11827233
Conc: 266.19 ng/ml



#29 AR-1254-4


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 57079660
Conc: 264.94 ng/ml



#30 AR-1254-5

R.T.: 7.170 min
Delta R.T.: 0.000 min
Response: 13869209
Conc: 265.15 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.231 min
Delta R.T.: 0.000 min
Response: 83258996
Conc: 263.21 ng/ml