

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090722\
 Data File : PR056334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2022 03:18
 Operator : AJ\MA
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 05:07:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 07 05:06:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.625	2.896	248.6E6	88843003	50.000	50.000
2) SA Decachlor...	9.513	8.104	99298338	82610052	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	3.847	3.118	26942080	9998911	500.000	500.000
9) L2 AR-1221-2	3.935	3.204	19593519	7553544	500.000	500.000
10) L2 AR-1221-3	4.013	3.279	59454621	23289777	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090722\
Data File : PR056334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2022 03:18
Operator : AJ\MA
Sample : AR1221ICC500
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 05:07:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 05:06:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.625 min
Delta R.T.: 0.000 min
Response: 248567870
Conc: 50.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 2.896 min
Delta R.T.: 0.000 min
Response: 88843003
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.513 min
Delta R.T.: 0.000 min
Response: 99298338
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.104 min
Delta R.T.: 0.000 min
Response: 82610052
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 3.847 min
Delta R.T.: 0.000 min
Response: 26942080
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.118 min
Delta R.T.: 0.000 min
Response: 9998911
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 3.935 min
Delta R.T.: 0.000 min
Response: 19593519
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.204 min
Delta R.T.: 0.000 min
Response: 7553544
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 59454621
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.279 min
Delta R.T.: 0.000 min
Response: 23289777
Conc: 500.00 ng/ml