

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090723\
 Data File : PR063077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2023 12:17
 Operator : YP\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12423252

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 15:40:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 29 05:38:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.679	2.950	45062851	60366467	19.554	19.915
2) SA Decachlor...	9.610	8.226	64218250	142.2E6	40.191	38.186
Target Compounds						
16) L4 AR-1242-1	4.907	4.072	25146747	34356137	359.656	354.520
17) L4 AR-1242-2	4.929	4.090	36967515	49296777	370.577	361.990
18) L4 AR-1242-3	4.994	4.270	22917162	26645408	366.233	359.739
19) L4 AR-1242-4	5.097	4.362	17933479	26445248	362.130	358.432
20) L4 AR-1242-5	5.888	4.910	18420498	34323162	361.769	365.985

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090723\
Data File : PR063077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2023 12:17
Operator : YP\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12423252

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 15:40:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 29 05:38:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

