

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR090920\  
Data File : PR047248.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 09 Sep 2020 10:10  
Operator : DD\AJ  
Sample : L3907-19DL 2X  
Misc :  
ALS Vial : 6 Sample Multiplier: 1


Instrument :  
ECD\_R  
ClientSampleId :  
ET1A6DL

Manual Integrations  
APPROVED

Ankita  
9/10/2020 1:10:50 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 10 01:28:38 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR090520CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Sep 05 06:49:21 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)  
R.T. Response Conc  
6.80 392138665 5973.54  
7.02 525231120 5147.80  
7.39 449704983 4124.68  
7.67 391423036 4619.36  
8.09 322460649 3411.33  
(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
5.80 1117699603 3929.80  
5.94 835993684 3422.90  
6.35 1828742037 3561.76  
6.58 1721691156 4427.24  
7.00 1732607477 3147.34

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR090920\  
Data File : PR047248.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 09 Sep 2020 10:10  
Operator : DD\AJ  
Sample : L3907-19DL 2X  
Misc :  
ALS Vial : 6 Sample Multiplier: 1



Instrument :  
ECD\_R  
ClientSampleId :  
ET1A6DL

Manual Integrations  
APPROVED

Ankita  
9/10/2020 1:10:50 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 10 01:28:38 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR090520CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Sep 05 06:49:21 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
6.80 392138665 5973.54  
7.02 525231120 5147.80  
7.39 449704983 4124.68  
7.67 391423036 4619.36  
8.09 322460649 3411.33  
(26) AR-1254-1 #2 (L6)  
R.T. Response Conc  
5.80 1117699603 3929.80  
5.94 835993684 3422.90  
6.35 1828742037 3561.76  
6.58 1561475315 4015.25  
7.00 1732607477 3147.34

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR090920\  
 Data File : PR047248.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 09 Sep 2020 10:10  
 Operator : DD\AJ  
 Sample : L3907-19DL 2X  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 ECD\_R  
 ClientSampleId :  
 ET1A6DL

Manual Integrations  
 APPROVED

Ankita  
 9/10/2020 1:10:50 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Sep 10 01:28:38 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR090520CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat Sep 05 06:49:21 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1   | Resp#2   | ng/ml     | ng/ml      |
|-----------------------------|--------|-------|----------|----------|-----------|------------|
| -----                       |        |       |          |          |           |            |
| System Monitoring Compounds |        |       |          |          |           |            |
| 1) SA Tetrachlo...          | 4.763  | 3.895 | 11689070 | 34503227 | 8.876     | 7.692      |
| 2) SA Decachlor...          | 10.718 | 8.989 | 32053681 | 243.5E6  | 19.120    | 18.820     |
| Target Compounds            |        |       |          |          |           |            |
| 16) L4 AR-1242-1            | 5.943  | 4.989 | 69999825 | 165.2E6  | 1531.089  | 1315.588   |
| 17) L4 AR-1242-2            | 5.965  | 5.009 | 220.9E6  | 649.2E6  | 3463.333  | 3050.208   |
| 18) L4 AR-1242-3            | 6.033  | 5.186 | 14239298 | 397.7E6  | 354.332   | 3131.336 # |
| 19) L4 AR-1242-4            | 6.129  | 5.269 | 110.1E6  | 551.1E6  | 3301.761  | 5889.714 # |
| 20) L4 AR-1242-5            | 6.867  | 5.796 | 481.5E6  | 1117.7E6 | 12466.507 | 7235.283 # |
| 26) L6 AR-1254-1            | 6.801  | 5.796 | 392.1E6  | 1117.7E6 | 5973.542  | 3929.804 # |
| 27) L6 AR-1254-2            | 7.022  | 5.944 | 525.2E6  | 836.0E6  | 5147.802  | 3422.904 # |
| 28) L6 AR-1254-3            | 7.389  | 6.352 | 449.7E6  | 1828.7E6 | 4124.683  | 3561.763   |
| 29) L6 AR-1254-4            | 7.675  | 6.580 | 391.4E6  | 1561.5E6 | 4619.356  | 4015.250m  |
| 30) L6 AR-1254-5            | 8.095  | 7.001 | 322.5E6  | 1732.6E6 | 3411.331  | 3147.343   |
| -----                       |        |       |          |          |           |            |

AJ  
 09/10/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.