

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091019CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 10 05:30:56 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040852.D 1600 =PR040854.D 800 =PR040853.D
 200 =PR040851.D 100 =PR040850.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.450	3.275	2.956	3.272	2.413	3.073 E6	13.35
2) SA	Decachlorobiphenyl	2.509	2.537	3.023	2.702	3.144	2.783 E6	10.32
3) L1	AR-1016-1	1.216	1.161	1.093	1.223	0.945	1.128 E5	10.18
4) L1	AR-1016-2	1.685	1.663	1.566	1.730	1.313	1.591 E5	10.47
5) L1	AR-1016-3	0.989	0.952	0.914	1.028	0.802	0.937 E5	9.25
6) L1	AR-1016-4	8.302	8.037	7.662	8.494	6.523	7.804 E4	10.01
7) L1	AR-1016-5	8.060	7.919	7.632	8.600	6.764	7.795 E4	8.66
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1	1.454	1.562	1.539	1.696	1.468	1.544 E5	6.26
32) L7	AR-1260-2	1.693	1.907	1.873	2.006	1.720	1.840 E5	7.15
33) L7	AR-1260-3	1.239	1.436	1.442	1.508	1.316	1.388 E5	7.82
34) L7	AR-1260-4	1.431	1.670	1.708	1.746	1.575	1.626 E5	7.75
35) L7	AR-1260-5	2.935	3.548	3.725	3.537	3.345	3.418 E5	8.82
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PR040852.D 1600 =PR040854.D 800 =PR040853.D
 200 =PR040851.D 100 =PR040850.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.534	1.285	1.247	1.487	1.106	1.332 E7	13.31
2) SA	Decachlorobiphenyl	0.887	0.862	1.030	0.963	1.145	0.977 E7	11.71
3) L1	AR-1016-1	2.890	2.544	2.465	3.003	2.333	2.647 E5	10.82
4) L1	AR-1016-2	3.978	3.567	3.466	4.127	3.213	3.670 E5	10.24
5) L1	AR-1016-3	2.200	1.990	1.940	2.280	1.749	2.032 E5	10.45
6) L1	AR-1016-4	1.768	1.563	1.545	1.849	1.441	1.633 E5	10.35

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091019CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Sep 10 05:30:56 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR040852.D 1600 =PR040854.D 800 =PR040853.D
 200 =PR040851.D 100 =PR040850.D

Compound			400	1600	800	200	100	Avg	%RSD
7)	L1	AR-1016-5	2.477	2.216	2.216	2.674	2.037	2.324 E5	10.79
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	4.848	4.907	4.994	5.748	4.976	5.094 E5	7.26
32)	L7	AR-1260-2	6.007	6.050	5.993	7.347	6.389	6.357 E5	9.07
33)	L7	AR-1260-3	5.391	5.761	5.817	6.370	5.498	5.768 E5	6.60
34)	L7	AR-1260-4	4.404	4.838	5.033	5.425	4.985	4.937 E5	7.47
35)	L7	AR-1260-5	1.150	1.294	1.384	1.384	1.292	1.301 E6	7.35
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range