

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091118\
 Data File : PR031959.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2018 09:04
 Operator : SM\SJ
 Sample : J4833-03 10X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ES-CABLE-OIL

Manual Integrations
 APPROVED

Sohil
 9/12/2018 9:13:37 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 09:37:14 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090618.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 07 03:46:12 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.525	3.656	47021146	85311427	2.398	2.314m
2) SA Decachlor...	10.278	8.523	54560230	31714103	1.696	1.211 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091118\
Data File : PR031959.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2018 09:04
Operator : SM\SJ
Sample : J4833-03 10X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
ES-CABLE-OIL


Manual Integrations
APPROVED

Sohil
9/12/2018 9:13:37 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 09:37:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090618.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 07 03:46:12 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.525 min
Delta R.T.: 0.000 min
Response: 47021146
Conc: 2.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
ES-CABLE-OIL


Manual Integrations
APPROVED

Sohil
9/12/2018 9:13:37 AM




#1 Tetrachloro-m-xylene

R.T.: 3.656 min
Delta R.T.: 0.003 min
Response: 85311427
Conc: 2.31 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.278 min
Delta R.T.: 0.006 min
Response: 54560230
Conc: 1.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.523 min
Delta R.T.: 0.005 min
Response: 31714103
Conc: 1.21 ng/ml