

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091119\
 Data File : PR040973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2019 16:28
 Operator : SM\AJ
 Sample : K4718-03
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AC6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 02:41:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:01:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.724	3.992	74786667	303.3E6	24.335	22.776
2)	SA Decachlor...	10.645	9.122	90087291	265.5E6	32.369	27.168
Target Compounds							
9)	L2 AR-1221-2	0.000	4.299	0	2674868	N.D.	34.833 #
28)	L6 AR-1254-3	0.000	6.471	0	85188334	N.D.	113.243 #
29)	L6 AR-1254-4	0.000	6.628f	0	6706501	N.D.	12.825 #
31)	L7 AR-1260-1	0.000	6.628f	0	6706501	N.D.	13.164 #
39)	L8 AR-1262-4	9.174f	0.000	21433807	0	105.480	N.D. #
42)	L9 AR-1268-2	9.174f	0.000	21433807	0	52.499	N.D. #
45)	L9 AR-1268-5	0.000	8.836	0	2484871	N.D.	0.674 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091119\
Data File : PR040973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2019 16:28
Operator : SM\AJ
Sample : K4718-03
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AC6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 02:41:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:01:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 5.020 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AC6




#9 AR-1221-2

R.T.: 4.299 min
Delta R.T.: 0.004 min
Response: 2674868
Conc: 34.83 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 7.346 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.471 min
Delta R.T.: 0.009 min
Response: 85188334
Conc: 113.24 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 7.631 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AC6




#29 AR-1254-4

R.T.: 6.628 min
Delta R.T.: -0.061 min
Response: 6706501
Conc: 12.83 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 7.507 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.628 min
Delta R.T.: 0.036 min
Response: 6706501
Conc: 13.16 ng/ml



#39 AR-1262-4


R.T.: 9.174 min
Delta R.T.: 0.042 min
Response: 21433807
Conc: 105.48 ng/ml

Instrument :
ECD_R
ClientSampleId :
C0AC6




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.002 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 9.174 min
Delta R.T.: 0.040 min
Response: 21433807
Conc: 52.50 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T. : 8.001 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 10.279 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AC6



#45 AR-1268-5

R.T.: 8.836 min
Delta R.T.: -0.003 min
Response: 2484871
Conc: 0.67 ng/ml