

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091218.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 13 01:38:59 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031992.D 1000 =PR031990.D 750 =PR031991.D
 250 =PR031993.D 50 =PR031994.D

Compound			500	1000	750	250	50	Avg		%RSD
1)	SA	Tetrachloro-m-xylene	2.185	2.051	2.235	2.109	2.300	2.176	E7	4.53
2)	SA	Decachlorobiphenyl	2.407	2.562	2.343	2.502	2.835	2.530	E7	7.52
3)	L1	AR-1016-1	5.638	5.348	5.770	5.521	6.461	5.748	E5	7.44
4)	L1	AR-1016-2	5.759	5.632	5.946	5.573	6.283	5.839	E5	4.91
5)	L1	AR-1016-3	0.870	0.861	0.917	0.847	1.001	0.899	E6	6.97
6)	L1	AR-1016-4	7.471	7.564	7.902	7.224	8.529	7.738	E5	6.52
7)	L1	AR-1016-5	6.300	6.557	6.794	6.103	7.162	6.583	E5	6.31
8)	L2	AR-1221-1	2.450					2.450	E5	0.00
9)	L2	AR-1221-2	1.815					1.815	E5	0.00
10)	L2	AR-1221-3	5.755					5.755	E5	0.00
11)	L3	AR-1232-1	5.204					5.204	E5	0.00
12)	L3	AR-1232-2	2.666					2.666	E5	0.00
13)	L3	AR-1232-3	6.989					6.989	E4	0.00
14)	L3	AR-1232-4	7.039					7.039	E4	0.00
15)	L3	AR-1232-5	4.900					4.900	E4	0.00
16)	L4	AR-1242-1	2.640					2.640	E5	0.00
17)	L4	AR-1242-2	4.744					4.744	E5	0.00
18)	L4	AR-1242-3	2.318					2.318	E5	0.00
19)	L4	AR-1242-4	1.730					1.730	E5	0.00
20)	L4	AR-1242-5	5.966					5.966	E5	0.00
21)	L5	AR-1248-1	7.735					7.735	E5	0.00
22)	L5	AR-1248-2	7.994					7.994	E5	0.00
23)	L5	AR-1248-3	1.073					1.073	E6	0.00
24)	L5	AR-1248-4	1.013					1.013	E6	0.00
25)	L5	AR-1248-5	6.700					6.700	E5	0.00
26)	L6	AR-1254-1	1.882					1.882	E6	0.00
27)	L6	AR-1254-2	1.173					1.173	E6	0.00
28)	L6	AR-1254-3	2.083					2.083	E6	0.00
29)	L6	AR-1254-4	1.636					1.636	E6	0.00
30)	L6	AR-1254-5	1.145					1.145	E6	0.00
31)	L7	AR-1260-1	1.221	1.375	1.358	1.203	1.410	1.313	E6	7.22
32)	L7	AR-1260-2	1.474	1.689	1.649	1.458	1.722	1.598	E6	7.74
33)	L7	AR-1260-3	1.590	1.892	1.790	1.567	1.781	1.724	E6	8.12
34)	L7	AR-1260-4	1.098	1.328	1.218	1.097	1.237	1.196	E6	8.26
35)	L7	AR-1260-5	2.496	3.095	2.761	2.463	2.699	2.703	E6	9.39
36)	L8	AR-1262-1	6.975					6.975	E5	0.00
37)	L8	AR-1262-2	5.976					5.976	E5	0.00
38)	L8	AR-1262-3	4.961					4.961	E5	0.00
39)	L8	AR-1262-4	1.394					1.394	E6	0.00
40)	L8	AR-1262-5	1.101					1.101	E6	0.00
41)	L9	AR-1268-1	6.915					6.915	E5	0.00
42)	L9	AR-1268-2	8.798					8.798	E5	0.00
43)	L9	AR-1268-3	3.707					3.707	E6	0.00
44)	L9	AR-1268-4	3.058					3.058	E6	0.00
45)	L9	AR-1268-5	9.477					9.477	E6	0.00

Signal #2 Calibration Files

500 =PR031992.D 1000 =PR031990.D 750 =PR031991.D
 250 =PR031993.D 50 =PR031994.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.201	3.917	4.322	4.057	4.482	4.196	E7	5.26
2) SA	Decachlorobiphenyl	1.779	1.824	1.647	1.912	2.302	1.893	E7	13.10
3) L1	AR-1016-1	1.157	1.029	1.166	1.145	1.377	1.175	E6	10.71
4) L1	AR-1016-2	1.756	1.655	1.804	1.667	2.007	1.778	E6	8.01
5) L1	AR-1016-3	3.059	2.878	3.231	3.030	3.680	3.176	E6	9.72

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091218.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 13 01:38:59 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031992.D 1000 =PR031990.D 750 =PR031991.D
 250 =PR031993.D 50 =PR031994.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.773	1.765	1.875	1.702	1.840	1.791	E6	3.78
7)	L1 AR-1016-5	1.634	1.621	1.752	1.575	1.963	1.709	E6	9.13
8)	L2 AR-1221-1	4.463					4.463	E5	0.00
9)	L2 AR-1221-2	3.297					3.297	E5	0.00
10)	L2 AR-1221-3	1.096					1.096	E6	0.00
11)	L3 AR-1232-1	5.735					5.735	E5	0.00
12)	L3 AR-1232-2	1.949					1.949	E5	0.00
13)	L3 AR-1232-3	5.726					5.726	E5	0.00
14)	L3 AR-1232-4	7.429					7.429	E5	0.00
15)	L3 AR-1232-5	9.470					9.470	E4	0.00
16)	L4 AR-1242-1	5.925					5.925	E5	0.00
17)	L4 AR-1242-2	1.229					1.229	E6	0.00
18)	L4 AR-1242-3	4.183					4.183	E5	0.00
19)	L4 AR-1242-4	3.923					3.923	E5	0.00
20)	L4 AR-1242-5	8.729					8.729	E5	0.00
21)	L5 AR-1248-1	1.708					1.708	E6	0.00
22)	L5 AR-1248-2	3.460					3.460	E6	0.00
23)	L5 AR-1248-3	2.468					2.468	E6	0.00
24)	L5 AR-1248-4	2.474					2.474	E6	0.00
25)	L5 AR-1248-5	1.596					1.596	E6	0.00
26)	L6 AR-1254-1	3.761					3.761	E6	0.00
27)	L6 AR-1254-2	2.876					2.876	E6	0.00
28)	L6 AR-1254-3	4.509					4.509	E6	0.00
29)	L6 AR-1254-4	2.171					2.171	E6	0.00
30)	L6 AR-1254-5	3.941					3.941	E6	0.00
31)	L7 AR-1260-1	3.189	3.588	3.489	3.244	3.860	3.474	E6	7.83
32)	L7 AR-1260-2	3.812	4.353	4.185	3.869	5.027	4.249	E6	11.50
33)	L7 AR-1260-3	3.045	3.479	3.331	3.082	3.788	3.345	E6	9.14
34)	L7 AR-1260-4	1.890	2.057	1.969	1.993	2.526	2.087	E6	12.10
35)	L7 AR-1260-5	3.961	4.243	4.033	4.220	5.586	4.409	E6	15.18
36)	L8 AR-1262-1	1.762					1.762	E6	0.00
37)	L8 AR-1262-2	1.363					1.363	E6	0.00
38)	L8 AR-1262-3	1.165					1.165	E6	0.00
39)	L8 AR-1262-4	1.488					1.488	E6	0.00
40)	L8 AR-1262-5	1.086					1.086	E6	0.00
41)	L9 AR-1268-1	1.536					1.536	E6	0.00
42)	L9 AR-1268-2	3.499					3.499	E6	0.00
43)	L9 AR-1268-3	2.825					2.825	E6	0.00
44)	L9 AR-1268-4	8.436					8.436	E5	0.00
45)	L9 AR-1268-5	7.084					7.084	E6	0.00

(#) = Out of Range