

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091219\
 Data File : PR041042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2019 14:01
 Operator : SM\AJ
 Sample : K4810-20
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 COAMO

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:45:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:01:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.724	3.993	53278536	221.5E6	17.336	16.633
2)	SA Decachlor...	10.646	9.120	92037093	290.0E6	33.070	29.668
Target Compounds							
8)	L2 AR-1221-1	0.000	4.260f	0	1086817	N.D.	9.058 #
9)	L2 AR-1221-2	0.000	4.298	0	2198789	N.D.	28.633 #
10)	L2 AR-1221-3	0.000	4.432f	0	2547328	N.D.	11.250 #
11)	L3 AR-1232-1	0.000	4.432f	0	2547328	N.D.	18.819 #
25)	L5 AR-1248-5	0.000	5.993f	0	4476939	N.D.	14.166 #
27)	L6 AR-1254-2	0.000	6.051	0	4845698	N.D.	10.974 #
28)	L6 AR-1254-3	0.000	6.464	0	123.6E6	N.D.	164.266 #
29)	L6 AR-1254-4	0.000	6.685	0	1388438	N.D.	2.655 #
30)	L6 AR-1254-5	0.000	7.103	0	1490576	N.D.	2.333 #
31)	L7 AR-1260-1	0.000	6.595	0	4214998	N.D.	8.274 #
33)	L7 AR-1260-3	0.000	6.934	0	2094994	N.D.	3.632 #
36)	L8 AR-1262-1	0.000	7.103f	0	1490576	N.D.	1.819 #
39)	L8 AR-1262-4	9.175f	0.000	33987640	0	167.260	N.D. #
42)	L9 AR-1268-2	9.175f	0.000	33987640	0	83.248	N.D. #
45)	L9 AR-1268-5	0.000	8.836	0	2029110	N.D.	0.550 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091219\
Data File : PR041042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2019 14:01
Operator : SM\AJ
Sample : K4810-20
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
COAM0

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:45:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:01:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 4.932 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#8 AR-1221-1

R.T.: 4.260 min
Delta R.T.: 0.052 min
Response: 1086817
Conc: 9.06 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.020 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.298 min
Delta R.T.: 0.003 min
Response: 2198789
Conc: 28.63 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 5.099 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#10 AR-1221-3

R.T.: 4.432 min
Delta R.T.: 0.059 min
Response: 2547328
Conc: 11.25 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 5.098 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.432 min
Delta R.T.: 0.059 min
Response: 2547328
Conc: 18.82 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.825 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#25 AR-1248-5

R.T.: 5.993 min
Delta R.T.: 0.046 min
Response: 4476939
Conc: 14.17 ng/ml




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T.: 6.976 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.051 min
Delta R.T.: -0.002 min
Response: 4845698
Conc: 10.97 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 7.346 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#28 AR-1254-3

R.T.: 6.464 min
Delta R.T.: 0.002 min
Response: 123571320
Conc: 164.27 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T.: 7.631 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.685 min
Delta R.T.: -0.004 min
Response: 1388438
Conc: 2.66 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 8.052 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#30 AR-1254-5

R.T.: 7.103 min
Delta R.T.: -0.007 min
Response: 1490576
Conc: 2.33 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T.: 7.507 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.595 min
Delta R.T.: 0.003 min
Response: 4214998
Conc: 8.27 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 8.123 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
COAM0




#33 AR-1260-3

R.T.: 6.934 min
Delta R.T.: -0.002 min
Response: 2094994
Conc: 3.63 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 8.125 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.103 min
Delta R.T.: -0.045 min
Response: 1490576
Conc: 1.82 ng/ml



#39 AR-1262-4


R.T.: 9.175 min
Delta R.T.: 0.042 min
Response: 33987640
Conc: 167.26 ng/ml

Instrument :
ECD_R
ClientSampleId :
COAM0




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.002 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 9.175 min
Delta R.T.: 0.041 min
Response: 33987640
Conc: 83.25 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T. : 8.001 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.279 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AM0



#45 AR-1268-5

R.T.: 8.836 min
Delta R.T.: -0.003 min
Response: 2029110
Conc: 0.55 ng/ml