

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091219\
 Data File : PR041050.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2019 16:40
 Operator : SM\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 111 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242317

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:49:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:01:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.723	3.992	53903954	227.2E6	17.540	17.064
2)	SA Decachlor...	10.643	9.118	90289620	279.8E6	32.442	28.632
Target Compounds							
16)	L4 AR-1242-1	5.900	5.093	29269863	64565496	393.923	371.631
17)	L4 AR-1242-2	5.924	5.113	40531455	88051309	386.734	367.910
18)	L4 AR-1242-3	5.986	5.292	24565276	50864381	388.115	374.644
19)	L4 AR-1242-4	6.086	5.375	20292369	49835451	384.019	362.415
20)	L4 AR-1242-5	6.823	5.901	23418320	75098279	371.650	327.445

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091219\
Data File : PR041050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2019 16:40
Operator : SM\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 111 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242317

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:49:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:01:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

