Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091222\
Data File : PR056589.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2022 23:18
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:14:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
	Target Compounds							
12) L3	AR-1232-2	4.555	0.000	550111	0	12.502	N.D.	#
39) L8	AR-1262-4	8.218	0.000	442033	0	6.423	N.D.	#
42) L9	AR-1268-2	8.218	0.000	442033	0	1.699	N.D.	#
43) L9	AR-1268-3	0.000	7.243	0	28260	N.D.	0.153	#
45) L9	AR-1268-5	9.201	0.000	265360	0	0.371	N.D.	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091222\
Data File : PR056589.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2022 23:18
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:14:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.624 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 2.896 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.512 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.103 min
Response: 0
Conc: N.D.



#12 AR-1232-2

R.T.: 4.555 min
Delta R.T.: -0.005 min
Response: 550111
Conc: 12.50 ng/ml

Instrument :
ECD_R
ClientSampleId :



#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.018 min
Response: 0
Conc: N.D.



#39 AR-1262-4

R.T.: 8.218 min
Delta R.T.: 0.012 min
Response: 442033
Conc: 6.42 ng/ml



#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.037 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.218 min
Delta R.T.: 0.009 min
Response: 442033
Conc: 1.70 ng/ml

Instrument :
ECD_R
ClientSampleId :



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.038 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.423 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 7.243 min
Delta R.T.: -0.013 min
Response: 28260
Conc: 0.15 ng/ml



#45 AR-1268-5

R.T.: 9.201 min
Delta R.T.: -0.005 min
Response: 265360
Conc: 0.37 ng/ml

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 7.866 min
Response: 0
Conc: N.D.