

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091222\
 Data File : PR056592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2022 00:05
 Operator : AJ\MA
 Sample : N4571-02
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 D2675

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/13/2022
 Supervised By :Ankita Jodhani 09/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 01:15:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 07 09:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.644f	2.893	89803265	39385319	18.691	22.471
2) SA Decachlor...	9.527	8.096	27074208	21360823	13.644m	12.839m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091222\
Data File : PR056592.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2022 00:05
Operator : AJ\MA
Sample : N4571-02
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
D2675


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/13/2022
Supervised By :Ankita Jodhani 09/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:15:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: 0.020 min
Response: 89803265
Conc: 18.69 ng/ml

Instrument :
ECD_R
ClientSampleId :
D2675


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/13/2022
Supervised By :Ankita Jodhani 09/13/2022




#1 Tetrachloro-m-xylene

R.T.: 2.893 min
Delta R.T.: -0.003 min
Response: 39385319
Conc: 22.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.527 min
Delta R.T.: 0.015 min
Response: 27074208
Conc: 13.64 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.096 min
Delta R.T.: -0.007 min
Response: 21360823
Conc: 12.84 ng/ml m