

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091322\
 Data File : PR056655.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 01:34
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 03:57:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 07 09:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.622	2.892	281.8E6	99468653	58.656	56.750
2) SA Decachlor...	9.504	8.094	116.2E6	89259186	58.575	53.649
Target Compounds						
26) L6 AR-1254-1	5.748	4.820	62842618	46987372	592.437	534.298
27) L6 AR-1254-2	5.984	4.978	92154029	40685802	590.123	530.228
28) L6 AR-1254-3	6.376	5.396	90380360	67221630	590.902	531.409
29) L6 AR-1254-4	6.688	5.640	67586390	42722004	581.190	537.444
30) L6 AR-1254-5	7.143	6.081	65539747	60174641	558.609	544.343



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091322\
Data File : PR056655.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 01:34
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 03:57:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.622 min
Delta R.T.: -0.002 min
Response: 281823913
Conc: 58.66 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.892 min
Delta R.T.: -0.004 min
Response: 99468653
Conc: 56.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.504 min
Delta R.T.: -0.008 min
Response: 116233008
Conc: 58.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.094 min
Delta R.T.: -0.010 min
Response: 89259186
Conc: 53.65 ng/ml



#26 AR-1254-1


R.T.: 5.748 min
Delta R.T.: -0.004 min
Response: 62842618
Conc: 592.44 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.820 min
Delta R.T.: -0.005 min
Response: 46987372
Conc: 534.30 ng/ml




#27 AR-1254-2

R.T.: 5.984 min
Delta R.T.: -0.004 min
Response: 92154029
Conc: 590.12 ng/ml



#27 AR-1254-2


R.T.: 4.978 min
Delta R.T.: -0.005 min
Response: 40685802
Conc: 530.23 ng/ml



#28 AR-1254-3


R.T.: 6.376 min
Delta R.T.: -0.005 min
Response: 90380360
Conc: 590.90 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.396 min
Delta R.T.: -0.005 min
Response: 67221630
Conc: 531.41 ng/ml




#29 AR-1254-4

R.T.: 6.688 min
Delta R.T.: -0.004 min
Response: 67586390
Conc: 581.19 ng/ml



#29 AR-1254-4


R.T.: 5.640 min
Delta R.T.: -0.005 min
Response: 42722004
Conc: 537.44 ng/ml



#30 AR-1254-5

R.T.: 7.143 min
Delta R.T.: -0.005 min
Response: 65539747
Conc: 558.61 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.081 min
Delta R.T.: -0.005 min
Response: 60174641
Conc: 544.34 ng/ml