

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091323\
 Data File : PR063298.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2023 05:33
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 ECD_R
 LabSampleId :
 AR1254CCC400

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 00:38:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 09/14/2023
 Supervised By :Ankita Jodhani 09/14/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
 R.T. Response Conc
 5.83 80046961 364.80
 6.07 120043198 362.25
 6.46 121992522 362.12
 6.78 81235587 350.52
 7.24 101396247 415.20

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 4.85 78170958 395.93
 5.01 68214524 403.56
 5.44 107272347 395.16
 5.68 65475394 390.28
 6.13 87444534 376.02

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091323\
Data File : PR063298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2023 05:33
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254CCC400

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 09/14/2023
Supervised By :Ankita Jodhani 09/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 06:51:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
5.83 80046961 364.80
6.07 120043198 362.25
6.46 121992522 362.12
6.78 81235587 350.52
7.24 87064369 356.51

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
4.85 78170958 395.93
5.01 68214524 403.56
5.44 107272347 395.16
5.68 65475394 390.28
6.13 87444534 376.02