

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091523\
 Data File : PR063335.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2023 12:03
 Operator : YP\AJ
 Sample : 04380-09
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 14:57:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.669	2.883	145.5E6	80461721	22.413	23.536
2)	SA Decachlor...	9.670	8.177	84734009	69631606	22.882	23.299
Target Compounds							
3)	L1 AR-1016-1	4.934	4.010	1453838	191768	7.442	1.704 #
4)	L1 AR-1016-2	4.934	4.028	1453838	143973	5.160	0.847 #
5)	L1 AR-1016-3	5.011	4.210	161399	390476	0.961	4.492 #
6)	L1 AR-1016-4	5.162f	4.262	242804	27079	1.769	0.400 #
7)	L1 AR-1016-5	5.407	4.472	600580	211365	4.500	2.441 #
8)	L2 AR-1221-1	3.880	3.119	918407	37284	12.653	1.007 #
9)	L2 AR-1221-2	3.986	3.187	1955536	916672	41.662	35.549
10)	L2 AR-1221-3	4.065	3.273	131960	94721	0.818	0.989
11)	L3 AR-1232-1	4.065	3.273	131960	94721	0.973	1.189
12)	L3 AR-1232-2	4.618	4.028	151232	143973	2.388	1.924
13)	L3 AR-1232-3	4.934	4.210	1453838	390476	11.556	10.371
14)	L3 AR-1232-4	5.162f	4.304	242804	33330	4.007	1.050 #
15)	L3 AR-1232-5	5.198	4.472	228906	211365	5.292	5.862
16)	L4 AR-1242-1	4.934	4.010	1453838	191768	8.721	2.009 #
17)	L4 AR-1242-2	4.934	4.028	1453838	143973	6.093	1.008 #
18)	L4 AR-1242-3	5.011	4.210	161399	390476	1.137	5.331 #
19)	L4 AR-1242-4	5.162f	4.304	242804	33330	2.072	0.487 #
20)	L4 AR-1242-5	5.845f	4.857	2005754	209754	17.297	2.398 #
21)	L5 AR-1248-1	4.934	4.010	1453838	191768	11.244	2.567 #
22)	L5 AR-1248-2	5.198	4.262	228906	27079	1.350	0.254 #
23)	L5 AR-1248-3	5.407	4.304	600580	33330	2.976	0.313 #
24)	L5 AR-1248-4	5.845	4.472	2005754	211365	9.206	1.627 #
25)	L5 AR-1248-5	5.845f	4.886	2005754	241092	9.533	1.876 #
26)	L6 AR-1254-1	5.845	4.857	2005754	209754	9.141	1.062 #
27)	L6 AR-1254-2	6.042	5.026	569194	1688226	1.718	9.988 #
28)	L6 AR-1254-3	6.451	5.444	1100945	382504	3.268	1.409 #
29)	L6 AR-1254-4	0.000	5.750f	0	4252829	N.D.	25.350 #
32)	L7 AR-1260-2	0.000	5.750f	0	4252829	N.D.	19.979 #
33)	L7 AR-1260-3	0.000	5.956	0	123813	N.D.	0.622 #
35)	L7 AR-1260-5	0.000	6.664f	0	4118471	N.D.	10.404 #
37)	L8 AR-1262-2	0.000	6.664f	0	4118471	N.D.	8.980 #
38)	L8 AR-1262-3	0.000	7.056	0	1117229	N.D.	6.279 #
39)	L8 AR-1262-4	0.000	7.056f	0	1117229	N.D.	3.321 #
41)	L9 AR-1268-1	0.000	7.056	0	1117229	N.D.	2.070 #
42)	L9 AR-1268-2	0.000	7.056f	0	1117229	N.D.	2.277 #
45)	L9 AR-1268-5	9.367	7.936	1209981	475908	0.788	0.355 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_R
ClientSampleId :

