

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091619\
 Data File : PR041161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2019 23:41
 Operator : SM\AJ
 Sample : K4845-13
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AN5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 01:34:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:01:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.723	3.991	4932690	24106266	1.605	1.810
2) SA Decachlor...	10.641	9.114	34123489	106.2E6	12.261	10.866
Target Compounds						
20) L4 AR-1242-5	0.000	5.944f	0	1831595	N.D.	7.986 #
25) L5 AR-1248-5	0.000	5.944	0	1831595	N.D.	5.796 #
26) L6 AR-1254-1	0.000	5.944f	0	1831595	N.D.	3.826 #
28) L6 AR-1254-3	0.000	6.454	0	51754144	N.D.	68.798 #
39) L8 AR-1262-4	9.170f	0.000	17795706	0	87.576	N.D. #
42) L9 AR-1268-2	9.170f	0.000	17795706	0	43.588	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091619\
Data File : PR041161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2019 23:41
Operator : SM\AJ
Sample : K4845-13
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AN5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 01:34:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:01:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.825 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AN5




#20 AR-1242-5

R.T.: 5.944 min
Delta R.T.: 0.039 min
Response: 1831595
Conc: 7.99 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.825 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.944 min
Delta R.T.: -0.003 min
Response: 1831595
Conc: 5.80 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.759 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AN5




#26 AR-1254-1

R.T.: 5.944 min
Delta R.T.: 0.038 min
Response: 1831595
Conc: 3.83 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 7.346 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.454 min
Delta R.T.: -0.008 min
Response: 51754144
Conc: 68.80 ng/ml



#39 AR-1262-4


R.T.: 9.170 min
Delta R.T.: 0.038 min
Response: 17795706
Conc: 87.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
C0AN5




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.002 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 9.170 min
Delta R.T.: 0.036 min
Response: 17795706
Conc: 43.59 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.001 min
Response: 0
Conc: N.D.