

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091619\
 Data File : PR041164.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Sep 2019 00:24
 Operator : SM\AJ
 Sample : K4845-16
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AN9

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 01:35:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 10 13:01:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.723	3.992	63548867	263.7E6	20.678	19.805
2) SA Decachlor...	10.642	9.115	87489184	271.6E6	31.436	27.791
Target Compounds						
9) L2 AR-1221-2	0.000	4.298	0	2099034	N.D.	27.334 #
20) L4 AR-1242-5	0.000	5.946f	0	4565127	N.D.	19.905 #
25) L5 AR-1248-5	0.000	5.946	0	4565127	N.D.	14.445 #
26) L6 AR-1254-1	0.000	5.946f	0	4565127	N.D.	9.535 #
29) L6 AR-1254-4	0.000	6.687	0	1876634	N.D.	3.589 #
45) L9 AR-1268-5	0.000	8.831	0	1714466	N.D.	0.465 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091619\
Data File : PR041164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Sep 2019 00:24
Operator : SM\AJ
Sample : K4845-16
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AN9

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 01:35:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 10 13:01:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 5.020 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AN9




#9 AR-1221-2

R.T.: 4.298 min
Delta R.T.: 0.003 min
Response: 2099034
Conc: 27.33 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.825 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.946 min
Delta R.T.: 0.041 min
Response: 4565127
Conc: 19.90 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.825 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AN9




#25 AR-1248-5

R.T.: 5.946 min
Delta R.T.: -0.001 min
Response: 4565127
Conc: 14.45 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.759 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.946 min
Delta R.T.: 0.039 min
Response: 4565127
Conc: 9.54 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.631 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
C0AN9




#29 AR-1254-4

R.T.: 6.687 min
Delta R.T.: -0.002 min
Response: 1876634
Conc: 3.59 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.279 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.831 min
Delta R.T.: -0.008 min
Response: 1714466
Conc: 0.46 ng/ml