

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091818.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 19 03:49:43 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR032100.D 1000 =PR032098.D 750 =PR032099.D
 250 =PR032101.D 50 =PR032102.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.919	1.902	2.041	1.953	2.125	1.988	E7	4.70
2) SA	Decachlorobiphenyl	3.051	2.585	2.695	2.896	3.419	2.929	E7	11.19
3) L1	AR-1016-1	4.978	4.643	5.055	5.288	6.033	5.199	E5	10.01
4) L1	AR-1016-2	5.040	4.685	5.101	5.355	6.189	5.274	E5	10.71
5) L1	AR-1016-3	7.760	7.331	7.888	8.229	9.658	8.173	E5	10.89
6) L1	AR-1016-4	6.922	6.361	6.872	7.253	8.389	7.159	E5	10.59
7) L1	AR-1016-5	5.990	5.379	5.792	6.220	7.143	6.105	E5	10.77
8) L2	AR-1221-1	2.681					2.681	E5	0.00
9) L2	AR-1221-2	2.045					2.045	E5	0.00
10) L2	AR-1221-3	6.238					6.238	E5	0.00
11) L3	AR-1232-1	4.830					4.830	E5	0.00
12) L3	AR-1232-2	2.481					2.481	E5	0.00
13) L3	AR-1232-3	1.129					1.129	E5	0.00
14) L3	AR-1232-4	7.544					7.544	E4	0.00
15) L3	AR-1232-5	6.519					6.519	E4	0.00
16) L4	AR-1242-1	2.237					2.237	E5	0.00
17) L4	AR-1242-2	4.182					4.182	E5	0.00
18) L4	AR-1242-3	2.194					2.194	E5	0.00
19) L4	AR-1242-4	1.690					1.690	E5	0.00
20) L4	AR-1242-5	5.541					5.541	E5	0.00
21) L5	AR-1248-1	1.252					1.252	E6	0.00
22) L5	AR-1248-2	1.303					1.303	E6	0.00
23) L5	AR-1248-3	1.773					1.773	E6	0.00
24) L5	AR-1248-4	1.693					1.693	E6	0.00
25) L5	AR-1248-5	1.118					1.118	E6	0.00
26) L6	AR-1254-1	1.511					1.511	E6	0.00
27) L6	AR-1254-2	9.841					9.841	E5	0.00
28) L6	AR-1254-3	1.782					1.782	E6	0.00
29) L6	AR-1254-4	1.404					1.404	E6	0.00
30) L6	AR-1254-5	9.570					9.570	E5	0.00
31) L7	AR-1260-1	1.365	1.144	1.219	1.317	1.488	1.307	E6	10.17
32) L7	AR-1260-2	1.749	1.461	1.543	1.661	1.864	1.656	E6	9.68
33) L7	AR-1260-3	2.084	1.718	1.817	1.922	2.050	1.918	E6	8.03
34) L7	AR-1260-4	1.507	1.214	1.281	1.388	1.512	1.380	E6	9.66
35) L7	AR-1260-5	3.347	2.795	2.860	2.978	3.316	3.059	E6	8.42
36) L8	AR-1262-1	7.598					7.598	E5	0.00
37) L8	AR-1262-2	7.074					7.074	E5	0.00
38) L8	AR-1262-3	6.078					6.078	E5	0.00
39) L8	AR-1262-4	1.813					1.813	E6	0.00
40) L8	AR-1262-5	1.354					1.354	E6	0.00
41) L9	AR-1268-1	7.082					7.082	E5	0.00
42) L9	AR-1268-2	9.246					9.246	E5	0.00
43) L9	AR-1268-3	4.262					4.262	E6	0.00
44) L9	AR-1268-4	3.431					3.431	E6	0.00
45) L9	AR-1268-5	1.018					1.018	E7	0.00

Signal #2 Calibration Files

500 =PR032100.D 1000 =PR032098.D 750 =PR032099.D
 250 =PR032101.D 50 =PR032102.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.024	4.001	4.251	4.104	4.542	4.185	E7	5.32
2) SA	Decachlorobiphenyl	2.258	1.918	2.043	2.238	2.787	2.249	E7	14.77
3) L1	AR-1016-1	1.099	1.033	1.101	1.198	1.354	1.157	E6	10.78
4) L1	AR-1016-2	1.658	1.544	1.650	1.722	2.024	1.720	E6	10.57
5) L1	AR-1016-3	3.062	2.744	3.047	3.228	3.725	3.161	E6	11.39

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR091818.M
 Title : GC EXTRACTABLES
 Last Update : Wed Sep 19 03:49:43 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR032100.D 1000 =PR032098.D 750 =PR032099.D
 250 =PR032101.D 50 =PR032102.D

Compound			500	1000	750	250	50	Avg	%RSD

6)	L1	AR-1016-4	1.758	1.676	1.740	1.793	1.983	1.790 E6	6.47
7)	L1	AR-1016-5	1.662	1.491	1.593	1.743	1.940	1.686 E6	10.06
8)	L2	AR-1221-1	5.313					5.313 E5	0.00
9)	L2	AR-1221-2	4.015					4.015 E5	0.00
10)	L2	AR-1221-3	1.253					1.253 E6	0.00
11)	L3	AR-1232-1	5.741					5.741 E5	0.00
12)	L3	AR-1232-2	1.925					1.925 E5	0.00
13)	L3	AR-1232-3	5.511					5.511 E5	0.00
14)	L3	AR-1232-4	7.035					7.035 E5	0.00
15)	L3	AR-1232-5	8.278					8.278 E4	0.00
16)	L4	AR-1242-1	5.383					5.383 E5	0.00
17)	L4	AR-1242-2	1.122					1.122 E6	0.00
18)	L4	AR-1242-3	3.973					3.973 E5	0.00
19)	L4	AR-1242-4	3.723					3.723 E5	0.00
20)	L4	AR-1242-5	8.375					8.375 E5	0.00
21)	L5	AR-1248-1	2.854					2.854 E6	0.00
22)	L5	AR-1248-2	5.712					5.712 E6	0.00
23)	L5	AR-1248-3	4.339					4.339 E6	0.00
24)	L5	AR-1248-4	3.958					3.958 E6	0.00
25)	L5	AR-1248-5	2.530					2.530 E6	0.00
26)	L6	AR-1254-1	3.128					3.128 E6	0.00
27)	L6	AR-1254-2	1.553					1.553 E6	0.00
28)	L6	AR-1254-3	3.222					3.222 E6	0.00
29)	L6	AR-1254-4	1.883					1.883 E6	0.00
30)	L6	AR-1254-5	2.202					2.202 E6	0.00
31)	L7	AR-1260-1	3.377	2.874	2.998	3.335	3.978	3.312 E6	12.97
32)	L7	AR-1260-2	3.944	3.521	3.565	4.084	5.323	4.087 E6	17.90
33)	L7	AR-1260-3	3.392	2.808	2.954	3.320	3.897	3.274 E6	12.99
34)	L7	AR-1260-4	2.070	1.610	1.762	2.074	2.510	2.005 E6	17.25
35)	L7	AR-1260-5	4.189	3.281	3.488	4.134	5.319	4.082 E6	19.52
36)	L8	AR-1262-1	1.730					1.730 E6	0.00
37)	L8	AR-1262-2	1.422					1.422 E6	0.00
38)	L8	AR-1262-3	1.017					1.017 E6	0.00
39)	L8	AR-1262-4	1.692					1.692 E6	0.00
40)	L8	AR-1262-5	1.054					1.054 E6	0.00
41)	L9	AR-1268-1	1.401					1.401 E6	0.00
42)	L9	AR-1268-2	5.994					5.994 E6	0.00
43)	L9	AR-1268-3	2.813					2.813 E6	0.00
44)	L9	AR-1268-4	8.000					8.000 E5	0.00
45)	L9	AR-1268-5	7.225					7.225 E6	0.00

 (#) = Out of Range