

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091918\
 Data File : PR032130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2018 15:48
 Operator : SM\SJ
 Sample : J4957-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 N48966

Manual Integrations
 APPROVED

Sohil
 9/20/2018 12:19:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 00:58:48 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 03:49:43 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.645	402.0E6	805.6E6	20.220	19.251m
2) SA Decachlor...	10.260	8.503	391.5E6	301.8E6	13.365	13.422

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091918\
Data File : PR032130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2018 15:48
Operator : SM\SJ
Sample : J4957-01
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
N48966


Manual Integrations
APPROVED

Sohil
9/20/2018 12:19:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 00:58:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 03:49:43 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.516 min
Delta R.T.: 0.002 min
Response: 401979753
Conc: 20.22 ng/ml

Instrument :
ECD_R
ClientSampleId :
N48966


Manual Integrations
APPROVED

Sohil
9/20/2018 12:19:29 PM




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: 0.001 min
Response: 805574698
Conc: 19.25 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: 0.005 min
Response: 391485559
Conc: 13.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.503 min
Delta R.T.: 0.000 min
Response: 301815164
Conc: 13.42 ng/ml