

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091922\
 Data File : PR056784.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2022 10:03
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 13:46:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 07 09:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.624	2.893	264.2E6	92557554	54.998	52.807
2)	SA Decachlor...	9.513	8.100	96395041	76820604	48.578	46.173
Target Compounds							
26)	L6 AR-1254-1	5.751	4.822	55710921	44270952	525.204	503.409
27)	L6 AR-1254-2	5.989	4.980	80383212	38094379	514.747	496.456
28)	L6 AR-1254-3	6.381	5.397	76304759	62103380	498.876	490.948
29)	L6 AR-1254-4	6.694	5.643	54688364	38852729	470.277	488.769
30)	L6 AR-1254-5	7.149	6.083	52254905	54787692	445.380	495.612



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091922\
Data File : PR056784.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2022 10:03
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 13:46:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 264247847
Conc: 55.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.893 min
Delta R.T.: -0.003 min
Response: 92557554
Conc: 52.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.513 min
Delta R.T.: 0.000 min
Response: 96395041
Conc: 48.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.100 min
Delta R.T.: -0.003 min
Response: 76820604
Conc: 46.17 ng/ml



#26 AR-1254-1


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 55710921
Conc: 525.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.822 min
Delta R.T.: -0.003 min
Response: 44270952
Conc: 503.41 ng/ml




#27 AR-1254-2

R.T.: 5.989 min
Delta R.T.: 0.000 min
Response: 80383212
Conc: 514.75 ng/ml



#27 AR-1254-2


R.T.: 4.980 min
Delta R.T.: -0.003 min
Response: 38094379
Conc: 496.46 ng/ml



#28 AR-1254-3


R.T.: 6.381 min
Delta R.T.: 0.000 min
Response: 76304759
Conc: 498.88 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.397 min
Delta R.T.: -0.003 min
Response: 62103380
Conc: 490.95 ng/ml




#29 AR-1254-4

R.T.: 6.694 min
Delta R.T.: 0.001 min
Response: 54688364
Conc: 470.28 ng/ml



#29 AR-1254-4


R.T.: 5.643 min
Delta R.T.: -0.003 min
Response: 38852729
Conc: 488.77 ng/ml



#30 AR-1254-5

R.T.: 7.149 min
Delta R.T.: 0.000 min
Response: 52254905
Conc: 445.38 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.083 min
Delta R.T.: -0.003 min
Response: 54787692
Conc: 495.61 ng/ml