

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091923\
 Data File : PR063474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 20:44
 Operator : YP\AJ
 Sample : 04417-06DL 5X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EZYF5DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 00:43:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.668	2.883	27541770	14810321	4.244	4.332
2) SA Decachlor...	9.671	8.177	17980201	12991769	4.855	4.347
Target Compounds						
31) L7 AR-1260-1	6.643	5.575	1989.9E6	1570.0E6	8282.670	8849.421
32) L7 AR-1260-2	6.929	5.783	2383.7E6	1923.4E6	8689.997	9035.442
33) L7 AR-1260-3	7.322	5.941	1430.7E6	1577.7E6	7481.867	7920.754
34) L7 AR-1260-4	7.565	6.452	1584.5E6	1261.9E6	7246.642	7706.475
35) L7 AR-1260-5	7.908	6.721	3182.9E6	2967.2E6	7354.672	7495.725

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR091923\
Data File : PR063474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 20:44
Operator : YP\AJ
Sample : 04417-06DL 5X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EZYF5DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 00:43:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

