

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092122\
 Data File : PR056911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2022 19:53
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 00:11:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 07 09:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.621	2.891	197.4E6	83060749	41.093	47.389
2) SA Decachlor...	9.505	8.093	103.2E6	82085432	52.024	49.337
Target Compounds						
26) L6 AR-1254-1	5.748	4.819	57295007	48143113	540.138	547.440
27) L6 AR-1254-2	5.985	4.977	80050920	39848425	512.619	519.316
28) L6 AR-1254-3	6.377	5.394	81233003	58965626	531.097	466.143
29) L6 AR-1254-4	6.689	5.639	53299964	36208808	458.338	455.508
30) L6 AR-1254-5	7.143	6.080	60604270	51894548	516.543	469.441



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092122\
Data File : PR056911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2022 19:53
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 00:11:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR090722.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 07 09:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.621 min
Delta R.T.: -0.003 min
Response: 197441043
Conc: 41.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.891 min
Delta R.T.: -0.005 min
Response: 83060749
Conc: 47.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.505 min
Delta R.T.: -0.007 min
Response: 103234784
Conc: 52.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.093 min
Delta R.T.: -0.010 min
Response: 82085432
Conc: 49.34 ng/ml



#26 AR-1254-1


R.T.: 5.748 min
Delta R.T.: -0.004 min
Response: 57295007
Conc: 540.14 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.819 min
Delta R.T.: -0.006 min
Response: 48143113
Conc: 547.44 ng/ml




#27 AR-1254-2

R.T.: 5.985 min
Delta R.T.: -0.004 min
Response: 80050920
Conc: 512.62 ng/ml



#27 AR-1254-2


R.T.: 4.977 min
Delta R.T.: -0.006 min
Response: 39848425
Conc: 519.32 ng/ml



#28 AR-1254-3


R.T.: 6.377 min
Delta R.T.: -0.004 min
Response: 81233003
Conc: 531.10 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.394 min
Delta R.T.: -0.007 min
Response: 58965626
Conc: 466.14 ng/ml




#29 AR-1254-4

R.T.: 6.689 min
Delta R.T.: -0.004 min
Response: 53299964
Conc: 458.34 ng/ml



#29 AR-1254-4


R.T.: 5.639 min
Delta R.T.: -0.007 min
Response: 36208808
Conc: 455.51 ng/ml



#30 AR-1254-5

R.T.: 7.143 min
Delta R.T.: -0.005 min
Response: 60604270
Conc: 516.54 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.080 min
Delta R.T.: -0.006 min
Response: 51894548
Conc: 469.44 ng/ml