

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR092319CLP.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 23 14:33:31 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR041388.D 1600 =PR041390.D 800 =PR041389.D
 200 =PR041387.D 100 =PR041386.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.225	1.167	1.228	1.273	1.132	1.205	E6	4.62
2) SA	Decachlorobiphenyl	3.457	3.339	3.422	3.703	3.501	3.484	E6	3.90
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	8.489	7.770	8.217	9.081	8.051	8.322	E4	5.99
27) L6	AR-1254-2	1.430	1.316	1.389	1.540	1.351	1.405	E5	6.14
28) L6	AR-1254-3	1.697	1.602	1.673	1.848	1.651	1.694	E5	5.48
29) L6	AR-1254-4	1.350	1.277	1.334	1.463	1.363	1.357	E5	4.96
30) L6	AR-1254-5	1.542	1.449	1.519	1.664	1.514	1.538	E5	5.11
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR041388.D 1600 =PR041390.D 800 =PR041389.D
 200 =PR041387.D 100 =PR041386.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.449	4.799	5.269	5.992	5.344	5.371	E6	7.95
2) SA	Decachlorobiphenyl	1.080	1.004	1.051	1.154	1.106	1.079	E7	5.23
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR092319CLP.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 23 14:33:31 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR041388.D 1600 =PR041390.D 800 =PR041389.D
 200 =PR041387.D 100 =PR041386.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.302	2.889	3.145	3.587	3.292	3.243 E5	7.84
27) L6	AR-1254-2	3.100	2.705	2.938	3.347	3.184	3.055 E5	8.02
28) L6	AR-1254-3	5.613	5.059	5.429	6.112	5.766	5.596 E5	6.99
29) L6	AR-1254-4	4.271	3.807	4.123	4.691	4.446	4.268 E5	7.80
30) L6	AR-1254-5	5.788	5.291	5.625	6.300	5.923	5.785 E5	6.43
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range