

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR092319CLP.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 23 16:13:43 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR041394.D 1600 =PR041396.D 800 =PR041395.D
 200 =PR041393.D 100 =PR041392.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.216	1.297	1.258	1.325	1.256	1.270	E6	3.28
2) SA	Decachlorobiphenyl	6.195	6.459	6.424	6.488	6.400	6.393	E6	1.81
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	4.609	4.945	4.851	4.861	4.778	4.809	E5	2.62
42) L9	AR-1268-2	4.257	4.571	4.473	4.467	4.347	4.423	E5	2.77
43) L9	AR-1268-3	3.548	3.813	3.740	3.715	3.660	3.695	E5	2.67
44) L9	AR-1268-4	1.713	1.751	1.752	1.799	1.739	1.751	E5	1.78
45) L9	AR-1268-5	1.379	1.475	1.455	1.402	1.362	1.414	E6	3.45

Signal #2 Calibration Files

400 =PR041394.D 1600 =PR041396.D 800 =PR041395.D
 200 =PR041393.D 100 =PR041392.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	5.346	5.171	5.305	6.179	6.031	5.606	E6	8.25
2) SA	Decachlorobiphenyl	1.885	1.889	1.924	1.979	1.973	1.930	E7	2.32
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR092319CLP.M
 Title : GC EXTRACTABLES
 Last Update : Mon Sep 23 16:13:43 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR041394.D 1600 =PR041396.D 800 =PR041395.D
 200 =PR041393.D 100 =PR041392.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.639	1.667	1.690	1.734	1.732	1.692 E6	2.45
42) L9	AR-1268-2	1.518	1.556	1.570	1.595	1.584	1.564 E6	1.91
43) L9	AR-1268-3	1.281	1.321	1.330	1.349	1.335	1.323 E6	1.95
44) L9	AR-1268-4	5.847	5.881	5.979	6.114	6.188	6.002 E5	2.45
45) L9	AR-1268-5	4.340	4.340	4.450	4.496	4.415	4.408 E6	1.56

(#) = Out of Range