

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092418\
 Data File : PR032253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2018 11:31
 Operator : SM\SJ
 Sample : J5051-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 092118-TIPC-F-D-S

Manual Integrations
 APPROVED

Sohil
 9/25/2018 12:01:12 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 04:11:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 03:49:43 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.648	290.9E6	508.4E6	14.632m	12.149
2) SA Decachlor...	10.250	8.497	376.1E6	283.3E6	12.839	12.597

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092418\
Data File : PR032253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2018 11:31
Operator : SM\SJ
Sample : J5051-08
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
092118-TIPC-F-D-S


Manual Integrations
APPROVED

Sohil
9/25/2018 12:01:12 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 04:11:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 03:49:43 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 290885283
Conc: 14.63 ng/ml m

Instrument :
ECD_R
ClientSampleId :
092118-TIPC-F-D-S


Manual Integrations
APPROVED

Sohil
9/25/2018 12:01:12 PM




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.003 min
Response: 508377423
Conc: 12.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: -0.005 min
Response: 376073643
Conc: 12.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.005 min
Response: 283266859
Conc: 12.60 ng/ml