

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092418\
 Data File : PR032270.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2018 05:08
 Operator : SM\SJ
 Sample : J4993-06 (Sig #1); J4993-06RE (Sig #2)
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 NS-OILY-WATER

Manual Integrations
 APPROVED

Sohil
 9/25/2018 12:01:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 07:35:13 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 19 03:49:43 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.646	259.7E6	564.6E6	13.063	13.493m
2) SA Decachlor...	10.250	8.498	325.3E6	246.8E6	11.107	10.973

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092418\
Data File : PR032270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2018 05:08
Operator : SM\SJ
Sample : J4993-06 (Sig #1); J4993-06RE (Sig #2)
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampled :
NS-OILY-WATER


Manual Integrations
APPROVED

Sohil
9/25/2018 12:01:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 07:35:13 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091818.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 19 03:49:43 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 259709620
Conc: 13.06 ng/ml

Instrument :
ECD_R
ClientSampleId :
NS-OILY-WATER


Manual Integrations
APPROVED

Sohil
9/25/2018 12:01:39 PM




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.002 min
Response: 564619616
Conc: 13.49 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: -0.005 min
Response: 325347370
Conc: 11.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.498 min
Delta R.T.: -0.004 min
Response: 246751156
Conc: 10.97 ng/ml