

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092622\
 Data File : PR056975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2022 23:24
 Operator : AJ\MA
 Sample : AR1262ICC800
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12624209

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/27/2022
 Supervised By :Ankita Jodhani 09/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 03:31:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 27 03:31:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.618	2.889	143.7E6	68781426	37.395	37.703
2) SA Decachlor...	9.503	8.092	148.0E6	120.8E6	75.783	72.966
Target Compounds						
36) L8 AR-1262-1	7.227	6.120	139.0E6	84997808	752.171	725.642
37) L8 AR-1262-2	7.810	6.658	225.2E6	163.4E6	759.246	751.574
38) L8 AR-1262-3	8.114	6.959	148.4E6	65775580	753.854	729.724
39) L8 AR-1262-4	8.200	7.028	72010855	125.0E6	760.584m	748.505
40) L8 AR-1262-5	8.819	7.564	78004397	59589108	745.949	733.532

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092622\
Data File : PR056975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2022 23:24
Operator : AJ\MA
Sample : AR1262ICC800
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12624209

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/27/2022
Supervised By :Ankita Jodhani 09/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 03:31:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 27 03:31:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

