

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092718\
 Data File : PR032372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2018 11:07
 Operator : SM\SJ
 Sample : J5078-01 (Sig #1); J5078-01RE (Sig #2)
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 600-WILSON-AVE-NEWARK

Manual Integrations
 APPROVED

Sohil
 9/28/2018 10:49:17 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 14:01:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092618.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 27 03:29:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.641	552.4E6	1106.4E6	25.903	24.624m
2) SA Decachlor...	10.234	8.493	600.1E6	455.8E6	20.549	20.347

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092718\
Data File : PR032372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2018 11:07
Operator : SM\SJ
Sample : J5078-01 (Sig #1); J5078-01RE (Sig #2)
Misc :
ALS Vial : 60 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
600-WILSON-AVE-NEWARK


Manual Integrations
APPROVED

Sohil
9/28/2018 10:49:17 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 14:01:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092618.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 27 03:29:23 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.508 min
Delta R.T.: 0.000 min
Response: 552359143
Conc: 25.90 ng/ml

Instrument :
ECD_R
ClientSampleId :
600-WILSON-AVE-NEWARK


Manual Integrations
APPROVED

Sohil
9/28/2018 10:49:17 AM




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 1106368887
Conc: 24.62 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: 0.002 min
Response: 600086671
Conc: 20.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.493 min
Delta R.T.: 0.000 min
Response: 455824320
Conc: 20.35 ng/ml