

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092918\
 Data File : PR032486.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2018 16:00
 Operator : SM\SJ
 Sample : J5125-06DL 20000X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0B74DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 04:19:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 28 17:37:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds								
31)	L7	AR-1260-1	7.263	6.128	1805.3E6	4789.2E6	1314.879	1304.234
32)	L7	AR-1260-2	7.517	6.313	2583.3E6	6704.8E6	1474.313	1437.376
33)	L7	AR-1260-3	7.876	6.463	1876.0E6	5085.4E6	1340.533	1347.237
34)	L7	AR-1260-4	8.102	6.923	2125.4E6	3124.1E6	1423.973	1292.541
35)	L7	AR-1260-5	8.423	7.163	5370.7E6	7391.5E6	1604.600	1488.315

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092918\
Data File : PR032486.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2018 16:00
Operator : SM\SJ
Sample : J5125-06DL 20000X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
C0B74DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 04:19:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 28 17:37:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

