

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092918\
 Data File : PR032482.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2018 15:02
 Operator : SM\SJ
 Sample : PB113342BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ALCS42

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 04:16:55 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Sep 28 17:37:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.646	321.7E6	654.7E6	16.180	18.849
2) SA Decachlor...	10.235	8.493	1289.8E6	968.9E6	38.809	36.847
Target Compounds						
3) L1 AR-1016-1	5.668	4.693	87064875	193.3E6	104.287	117.598
4) L1 AR-1016-2	5.690	4.709	122.5E6	340.3E6	99.705	105.739
5) L1 AR-1016-3	5.753	4.882	75327230	157.4E6	100.719	102.063
6) L1 AR-1016-4	5.850	4.920	62174994	142.4E6	98.948	112.969
7) L1 AR-1016-5	6.143	5.128	64877766	188.0E6	102.234	108.453
31) L7 AR-1260-1	7.263	6.128	143.7E6	409.1E6	104.693	111.412
32) L7 AR-1260-2	7.516	6.313	185.9E6	540.9E6	106.086	115.955
33) L7 AR-1260-3	7.875	6.462	134.0E6	428.3E6	95.730	113.463
34) L7 AR-1260-4	8.101	6.923	152.9E6	265.2E6	102.467	109.720
35) L7 AR-1260-5	8.421	7.162	344.5E6	597.7E6	102.930	120.344

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092918\
Data File : PR032482.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2018 15:02
Operator : SM\SJ
Sample : PB113342BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ALCS42

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 01 04:16:55 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Sep 28 17:37:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

