

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
 Data File : PR057038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Sep 2022 17:57
 Operator : AJ\MA
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 04:08:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 29 04:08:13 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.616	2.887	165.2E6	91214611	50.000	50.000
2) SA Decachlor...	9.500	8.089	106.8E6	82317781	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	4.837	3.990	50223747	25287097	500.000	500.000
4) L1 AR-1016-2	4.859	4.007	74230747	39223338	500.000	500.000
5) L1 AR-1016-3	4.923	4.184	45472326	21379218	500.000	500.000
6) L1 AR-1016-4	5.026	4.235	36731070	15606491	500.000	500.000
7) L1 AR-1016-5	5.341	4.449	35996925	20488054	500.000	500.000
31) L7 AR-1260-1	6.552	5.528	64474800	38397192	500.000	500.000
32) L7 AR-1260-2	6.835	5.732	78726734	46605979	500.000	500.000
33) L7 AR-1260-3	7.226	5.888	62261516	43648253	500.000	500.000
34) L7 AR-1260-4	7.468	6.390	63586115	36877770	500.000	500.000
35) L7 AR-1260-5	7.807	6.655	128.8E6	88354328	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
Data File : PR057038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Sep 2022 17:57
Operator : AJ\MA
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 04:08:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 29 04:08:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.616 min
Delta R.T.: 0.000 min
Response: 165189125
Conc: 50.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 2.887 min
Delta R.T.: 0.000 min
Response: 91214611
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.500 min
Delta R.T.: 0.000 min
Response: 106767139
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.089 min
Delta R.T.: 0.000 min
Response: 82317781
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 4.837 min
Delta R.T.: 0.000 min
Response: 50223747
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 25287097
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 4.859 min
Delta R.T.: 0.000 min
Response: 74230747
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 39223338
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 45472326
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.184 min
Delta R.T.: 0.000 min
Response: 21379218
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.026 min
Delta R.T.: 0.000 min
Response: 36731070
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.235 min
Delta R.T.: 0.000 min
Response: 15606491
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.341 min
Delta R.T.: 0.000 min
Response: 35996925
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 4.449 min
Delta R.T.: 0.000 min
Response: 20488054
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.552 min
Delta R.T.: 0.000 min
Response: 64474800
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 5.528 min
Delta R.T.: 0.000 min
Response: 38397192
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 6.835 min
Delta R.T.: 0.000 min
Response: 78726734
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 5.732 min
Delta R.T.: 0.000 min
Response: 46605979
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.226 min
Delta R.T.: 0.000 min
Response: 62261516
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 5.888 min
Delta R.T.: 0.000 min
Response: 43648253
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.468 min
Delta R.T.: 0.000 min
Response: 63586115
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.390 min
Delta R.T.: 0.000 min
Response: 36877770
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.807 min
Delta R.T.: 0.000 min
Response: 128827466
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 6.655 min
Delta R.T.: 0.000 min
Response: 88354328
Conc: 500.00 ng/ml