

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
 Data File : PR057064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2022 00:45
 Operator : AJ\MA
 Sample : PR092922ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR092922

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 11:32:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 29 10:58:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.617	2.887	149.2E6	80585369	46.415	46.344
2) SA Decachlor...	9.496	8.086	105.7E6	81387796	51.731	50.433
Target Compounds						
3) L1 AR-1016-1	4.838	3.990	45634868	22479643	462.587	457.391
4) L1 AR-1016-2	4.860	4.006	67763080	35589324	470.421	475.634
5) L1 AR-1016-3	4.924	4.183	41907688	19268853	475.723	473.229
6) L1 AR-1016-4	5.026	4.234	33842120	14034502	478.548	468.717
7) L1 AR-1016-5	5.340	4.448	32936001	18595720	467.920	472.394
31) L7 AR-1260-1	6.551	5.527	59505343	35308484	470.676	472.930
32) L7 AR-1260-2	6.835	5.732	72592337	43373990	470.873	474.398
33) L7 AR-1260-3	7.225	5.887	57880012	40682390	471.709	481.288
34) L7 AR-1260-4	7.467	6.389	60085013	34749825	481.443	485.121
35) L7 AR-1260-5	7.805	6.654	121.3E6	84313176	484.489	488.087



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
Data File : PR057064.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2022 00:45
Operator : AJ\MA
Sample : PR092922ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR092922

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 11:32:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 29 10:58:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.617 min
Delta R.T.: 0.000 min
Response: 149198621
Conc: 46.41 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#1 Tetrachloro-m-xylene

R.T.: 2.887 min
Delta R.T.: 0.000 min
Response: 80585369
Conc: 46.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.496 min
Delta R.T.: -0.001 min
Response: 105676321
Conc: 51.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.086 min
Delta R.T.: -0.003 min
Response: 81387796
Conc: 50.43 ng/ml



#3 AR-1016-1


R.T.: 4.838 min
Delta R.T.: 0.000 min
Response: 45634868
Conc: 462.59 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#3 AR-1016-1

R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 22479643
Conc: 457.39 ng/ml




#4 AR-1016-2

R.T.: 4.860 min
Delta R.T.: 0.000 min
Response: 67763080
Conc: 470.42 ng/ml



#4 AR-1016-2


R.T.: 4.006 min
Delta R.T.: 0.000 min
Response: 35589324
Conc: 475.63 ng/ml



#5 AR-1016-3


R.T.: 4.924 min
Delta R.T.: 0.000 min
Response: 41907688
Conc: 475.72 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#5 AR-1016-3

R.T.: 4.183 min
Delta R.T.: 0.000 min
Response: 19268853
Conc: 473.23 ng/ml




#6 AR-1016-4

R.T.: 5.026 min
Delta R.T.: 0.000 min
Response: 33842120
Conc: 478.55 ng/ml



#6 AR-1016-4


R.T.: 4.234 min
Delta R.T.: 0.000 min
Response: 14034502
Conc: 468.72 ng/ml



#7 AR-1016-5


R.T.: 5.340 min
Delta R.T.: 0.000 min
Response: 32936001
Conc: 467.92 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#7 AR-1016-5

R.T.: 4.448 min
Delta R.T.: 0.000 min
Response: 18595720
Conc: 472.39 ng/ml




#31 AR-1260-1

R.T.: 6.551 min
Delta R.T.: 0.000 min
Response: 59505343
Conc: 470.68 ng/ml



#31 AR-1260-1


R.T.: 5.527 min
Delta R.T.: 0.000 min
Response: 35308484
Conc: 472.93 ng/ml



#32 AR-1260-2


R.T.: 6.835 min
Delta R.T.: 0.000 min
Response: 72592337
Conc: 470.87 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#32 AR-1260-2

R.T.: 5.732 min
Delta R.T.: 0.000 min
Response: 43373990
Conc: 474.40 ng/ml




#33 AR-1260-3

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 57880012
Conc: 471.71 ng/ml



#33 AR-1260-3


R.T.: 5.887 min
Delta R.T.: 0.000 min
Response: 40682390
Conc: 481.29 ng/ml



#34 AR-1260-4


R.T.: 7.467 min
Delta R.T.: 0.000 min
Response: 60085013
Conc: 481.44 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922




#34 AR-1260-4

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 34749825
Conc: 485.12 ng/ml



#35 AR-1260-5

R.T.: 7.805 min
Delta R.T.: -0.002 min
Response: 121283233
Conc: 484.49 ng/ml



#35 AR-1260-5

R.T.: 6.654 min
Delta R.T.: -0.001 min
Response: 84313176
Conc: 488.09 ng/ml