

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
 Data File : PR057066.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2022 01:16
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR092922AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 11:33:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 29 10:58:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.616	2.887	159.7E6	87769181	49.681	50.476
2) SA Decachlor...	9.494	8.085	110.4E6	84661662	54.065	52.462
Target Compounds						
21) L5 AR-1248-1	4.837	3.990	30580317	15013905	491.351	492.608
22) L5 AR-1248-2	5.128	4.234	39612971	20858173	489.460	494.405
23) L5 AR-1248-3	5.340	4.274	47193670	21730304	487.170	493.664
24) L5 AR-1248-4	5.773	4.448	51639965	26492806	494.806	494.112
25) L5 AR-1248-5	5.812	4.856	49641892	26667422	496.289	501.761



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092922\
Data File : PR057066.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2022 01:16
Operator : AJ\MA
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR092922AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 11:33:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR092922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 29 10:58:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.616 min
Delta R.T.: -0.001 min
Response: 159696803
Conc: 49.68 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922AR1248




#1 Tetrachloro-m-xylene

R.T.: 2.887 min
Delta R.T.: 0.000 min
Response: 87769181
Conc: 50.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.494 min
Delta R.T.: -0.004 min
Response: 110446082
Conc: 54.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.085 min
Delta R.T.: -0.004 min
Response: 84661662
Conc: 52.46 ng/ml



#21 AR-1248-1


R.T.: 4.837 min
Delta R.T.: 0.000 min
Response: 30580317
Conc: 491.35 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922AR1248




#21 AR-1248-1

R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 15013905
Conc: 492.61 ng/ml




#22 AR-1248-2

R.T.: 5.128 min
Delta R.T.: 0.000 min
Response: 39612971
Conc: 489.46 ng/ml



#22 AR-1248-2


R.T.: 4.234 min
Delta R.T.: 0.000 min
Response: 20858173
Conc: 494.41 ng/ml



#23 AR-1248-3


R.T.: 5.340 min
Delta R.T.: 0.000 min
Response: 47193670
Conc: 487.17 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922AR1248




#23 AR-1248-3

R.T.: 4.274 min
Delta R.T.: 0.000 min
Response: 21730304
Conc: 493.66 ng/ml




#24 AR-1248-4

R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 51639965
Conc: 494.81 ng/ml



#24 AR-1248-4


R.T.: 4.448 min
Delta R.T.: 0.000 min
Response: 26492806
Conc: 494.11 ng/ml



#25 AR-1248-5

R.T.: 5.812 min
Delta R.T.: 0.000 min
Response: 49641892
Conc: 496.29 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR092922AR1248



#25 AR-1248-5

R.T.: 4.856 min
Delta R.T.: 0.000 min
Response: 26667422
Conc: 501.76 ng/ml