Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
Data File : PR063700.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 09:49
Operator : YP\AJ
Sample : PB155922BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 14:20:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	3.665	2.880	119.5E6	67888551	18.419	19.858	
2)	SA Decachlor...	9.663	8.171	78488080	64042980	21.195	21.429	
	Target Compounds							
3)	L1 AR-1016-1	4.932	4.016	739639	120567	3.786	1.071	#
4)	L1 AR-1016-2	4.932	4.039	739639	350909	2.625	2.065	
5)	L1 AR-1016-3	4.989	4.217	114768	146913	0.683	1.690	#
6)	L1 AR-1016-4	0.000	4.268	0	338859	N.D.	5.010	#
7)	L1 AR-1016-5	0.000	4.463	0	278496	N.D.	3.216	#
8)	L2 AR-1221-1	3.899	3.099	780888	73725	10.759	1.991	#
9)	L2 AR-1221-2	3.982	3.185	1898793	476944	40.453	18.496	#
10)	L2 AR-1221-3	4.036	3.277	1225391	742838	7.595	7.753	
11)	L3 AR-1232-1	4.036	3.277	1225391	742838	9.034	9.321	
12)	L3 AR-1232-2	0.000	4.039	0	350909	N.D.	4.690	#
13)	L3 AR-1232-3	4.932	4.217	739639	146913	5.879	3.902	#
14)	L3 AR-1232-4	0.000	4.323	0	374799	N.D.	11.813	#
15)	L3 AR-1232-5	5.212	4.463	57543	278496	1.330	7.724	#
16)	L4 AR-1242-1	4.932	4.016	739639	120567	4.437	1.263	#
17)	L4 AR-1242-2	4.932	4.039	739639	350909	3.100	2.457	
18)	L4 AR-1242-3	4.989	4.217	114768	146913	0.808	2.006	#
19)	L4 AR-1242-4	0.000	4.323	0	374799	N.D.	5.480	#
20)	L4 AR-1242-5	5.900	4.846	719781	357494	6.207	4.087	#
21)	L5 AR-1248-1	4.932	4.016	739639	120567	5.720	1.614	#
22)	L5 AR-1248-2	5.212	4.254	57543	157941	0.339	1.480	#
23)	L5 AR-1248-3	0.000	4.323	0	374799	N.D.	3.524	#
24)	L5 AR-1248-4	5.828	4.463	1673871	278496	7.683	2.144	#
25)	L5 AR-1248-5	5.900	4.868	719781	838849	3.421	6.526	#
26)	L6 AR-1254-1	5.828	4.846	1673871	357494	7.628	1.811	#
27)	L6 AR-1254-2	6.109f	5.009	967567	421938	2.920	2.496	
28)	L6 AR-1254-3	6.446	5.442	2756731	709136	8.183	2.612	#
29)	L6 AR-1254-4	6.756	5.681	534485	153472	2.306	0.915	#
30)	L6 AR-1254-5	7.216	6.123	1115417	304228	4.567	1.308	#
31)	L7 AR-1260-1	0.000	5.590	0	302558	N.D.	1.705	#
32)	L7 AR-1260-2	0.000	5.782	0	196478	N.D.	0.923	#
33)	L7 AR-1260-3	0.000	5.947	0	149235	N.D.	0.749	#
34)	L7 AR-1260-4	0.000	6.458	0	66776	N.D.	0.408	#
36)	L8 AR-1262-1	0.000	6.211	0	488542	N.D.	1.932	#
38)	L8 AR-1262-3	0.000	7.043	0	605775	N.D.	3.405	#
39)	L8 AR-1262-4	0.000	7.043f	0	605775	N.D.	1.801	#
41)	L9 AR-1268-1	0.000	7.043	0	605775	N.D.	1.122	#
42)	L9 AR-1268-2	0.000	7.043f	0	605775	N.D.	1.234	#
43)	L9 AR-1268-3	0.000	7.316	0	177047	N.D.	0.426	#
45)	L9 AR-1268-5	9.347	7.934	814621	424254	0.530	0.316	#

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
Data File : PR063700.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 09:49
Operator : YP\AJ
Sample : PB155922BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 14:20:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
Data File : PR063700.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 09:49
Operator : YP\AJ
Sample : PB155922BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 14:20:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm