

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
 Data File : PR063718.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2023 14:15
 Operator : YP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12483373

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 21:15:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.666	2.880	128.6E6	72576195	19.822	21.229
2)	SA Decachlor...	9.664	8.170	135.9E6	102.3E6	36.699	34.232
Target Compounds							
21)	L5 AR-1248-1	4.905	4.004	50241937	28833319	388.567	386.023
22)	L5 AR-1248-2	5.201	4.253	68363918	42449322	403.140	397.654
23)	L5 AR-1248-3	5.415	4.294	77881155	41877156	385.909	393.758
24)	L5 AR-1248-4	5.852	4.471	79434112	49943847	364.582	384.434
25)	L5 AR-1248-5	5.892	4.886	77205977	49887388	366.928	388.102

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
Data File : PR063718.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 14:15
Operator : YP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12483373

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 21:15:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

