

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
 Data File : PR063711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2023 12:33
 Operator : YP\AJ
 Sample : 04527-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EZYY8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/02/2023
 Supervised By :Ankita Jodhani 10/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 14:24:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.666	2.880	138.2E6	81261574	21.293	23.770
2)	SA Decachlor...	9.665	8.170	77717384	58127490	20.987	19.450
Target Compounds							
16)	L4 AR-1242-1	4.905	4.004	20139543	11396564	120.803	119.378m
17)	L4 AR-1242-2	4.927	4.021	30371920	15595517	127.285	109.215m
18)	L4 AR-1242-3	4.992	4.201	15690345	10531770	110.506	143.773m#
19)	L4 AR-1242-4	5.095	4.294	15807126	14652717	134.869	214.231m#
20)	L4 AR-1242-5	5.891	4.843	31302748	32457604	269.942	371.047m#
31)	L7 AR-1260-1	6.639	5.569	54708424	45585113	227.715	256.948m
32)	L7 AR-1260-2	6.924	5.777	72069669	59836218	262.741	281.095m
33)	L7 AR-1260-3	7.317	5.935	47748734	41154572	249.704	206.619m
34)	L7 AR-1260-4	7.561	6.445	43625187	33809140	199.513m	206.471
35)	L7 AR-1260-5	7.904	6.715	92952566	80016688	214.785	202.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR092923\
Data File : PR063711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2023 12:33
Operator : YP\AJ
Sample : 04527-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EZY8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/02/2023
Supervised By :Ankita Jodhani 10/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 14:24:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

