

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100223\
 Data File : PR063787.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2023 11:05
 Operator : YP\AJ
 Sample : 04469-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EZYL8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/03/2023
 Supervised By :Ankita Jodhani 10/03/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 22:17:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 13 07:34:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.665	2.879	131.3E6	74722497	20.227	21.857
2) SA Decachlor...	9.665	8.172	83509048	61992298	22.551	20.743
Target Compounds						
31) L7 AR-1260-1	6.638	5.568	23558992	17023850	98.061	95.958m
32) L7 AR-1260-2	6.924	5.777	25990283	22632715	94.752	106.323
33) L7 AR-1260-3	7.319	5.934	15578717	15365599	81.470	77.144
34) L7 AR-1260-4	7.561	6.444	17444728	11081138	79.781m	67.672
35) L7 AR-1260-5	7.904	6.714	33452903	26549335	77.299	67.069

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR100223\
Data File : PR063787.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2023 11:05
Operator : YP\AJ
Sample : 04469-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EZYL8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/03/2023
Supervised By :Ankita Jodhani 10/03/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 22:17:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR091223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 13 07:34:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

